

Consona N9 series

Diagnostic Ultrasound System

Datasheet

Release 02(02.01.00)

Concrete to Resonate

1 System Overview	
	<p>Dedicatedly design for Primary Care, Consona Series feature progressive and concrete solution. No matter in hospital, clinic or imaging center, the user of Consona will experience the powerful performance, smart application and the ergonomic design, and obtain the values of confident diagnosis, convenient operation and conducive improvement.</p>
1.1	Advantages of ZST+ platform <ul style="list-style-type: none"> Advanced Acoustic Acquisition Dynamic Pixel Focusing (DPF) Total Recall Imaging (TRI) Powerful Processing Architecture Enhanced Channel Data Processing ZONE Sonography® Technology Up to 248,832 channels
1.2	Application <ul style="list-style-type: none"> Abdomen Obstetrics Gynecology Cardiology Small parts Urology Vascular Pediatrics Nerve Emergency & Critical Pelvic Floor
1.3	Transducer types <ul style="list-style-type: none"> Curved array transducer Linear array transducer Phased array transducer 4D Volume transducer Pencil transducer
1.4	Advanced Imaging technics <ul style="list-style-type: none"> THI (Tissue Harmonic Imaging) and PSH (Phase Shift Harmonic Imaging) iBeam (Spatial Compound Imaging) iClear (Speckle Suppression Imaging) iTouch⁺ (Auto Image Optimization) Echo Boost Zoom/iZoom (Full Screen Zoom) FCI (Frequency Compound Imaging) B steer
1.5	<ul style="list-style-type: none"> HD Scope Smart Track HR Flow (High Resolution Flow) Glazing Flow Imaging modes <ul style="list-style-type: none"> B-Mode M-Mode/Color M-mode Color Doppler Imaging Power Doppler Imaging/Directional PDI Pulsed Wave Doppler Continuous Wave Doppler Free Xros M (Anatomical M-mode) Free Xros CM (Curved Anatomical M-mode)
1.6	Function modes <ul style="list-style-type: none"> CW 4D iScape View Free Xros M Free Xros CM Tissue Doppler Imaging TDI QA Contrast imaging Contrast Imaging QA VLMI Contrast Volume CEUS Strain Elastography STE STQ Stress Echo Tissue Tracking QA Smart 3D iPage⁺ SCV⁺ iLive STIC Color 3D Niche Smart Volume Smart V Trace Smart Planes CNS Smart Face Smart Scene 3D Glazing Flow iNeedle Abdomen/General Package Obstetrics Package

		<ul style="list-style-type: none"> • Gynecology Package • Cardiology Package • Small Parts Package • Urology Package • Vascular Package • Pediatrics Package • Nerve Package • Emergency&Critical Package • Smart Pelvic • Smart OB • Smart NT • Smart Fetal HR • Smart FLC • IVF • IMT • RIMT • AutoEF • R-VQS • Smart Thyroid • Smart Breast • Smart Hip • Smart HRI • Smart Bladder • Smart Trace • CPP • Smart B-line • V-Mapping • DICOM Basic • DICOM Worklist • DICOM MPPS • DICOM Query/Retrieve • DICOM OB/GYN SR • DICOM Vascular SR • DICOM Cardiac SR • DICOM Breast SR • DICOM Abdomen SR • DICOM Small Parts SR • iWorks • DVR Module • iVocal • ClamAV • Smart Calc • Smart Echovue • Smart ABDvue
1.7	1.8	<p>Language support</p> <ul style="list-style-type: none"> • Software: English, Chinese, German, Spanish, French, Italian, Portuguese, Russian, Czech, Polish, Turkish, Finnish, Danish, Icelandic, Norwegian, Swedish, Hungarian, Serbian, Dutch, Lithuanian, Greek, Thai • Keyboard input: English, Chinese, German, Spanish, French, Italian, Portuguese, Russian, Czech, Polish, Icelandic, Norwegian, Swedish, Finnish, Turkish, Danish, Hungarian, Serbian • User manual: English, Chinese, German, Spanish, French, Italian, Portuguese, Russian, Polish, Turkish, Serbian, Dutch, Greek • QWERTY Keyboard: English, Russian
	2	<p>Physical Specification</p>
	2.1	<p>Dimension and weight</p> <ul style="list-style-type: none"> • The control panel and the monitor is in the lowest position. Configured with dual-wing floating support arm and 23.8-inch monitor <ul style="list-style-type: none"> – Depth: 840±40mm – Width: 542±10mm – Height: 988±20mm • The control panel and the monitor is in the highest position. Configured with dual-wing floating support arm and 23.8-inch monitor
1.7	Other features	<ul style="list-style-type: none"> • Multilingual control panel overlay • GPU

	<ul style="list-style-type: none"> - Depth: 740±40mm - Width: 542±10mm - Height: 1690±20mm • Weight: ≤65kg (net weight, standard configuration) 	3.2	<ul style="list-style-type: none"> • 3 keys for user-defined functions • Rotate angle range: 180 degrees • Down/up: 300±20 mm • Front/Rear: 110 mm <p>Monitor</p> <ul style="list-style-type: none"> • 23.8-inch bezel-less LED monitor with high resolution • Resolution: 1920x1080 • Viewing angle: 178 degrees • Digital on screen display of brightness and contrast controls • Automatic adjustment of monitor light with the changing environment • Automatic LED brightness • Tilt/Rotate independent adjustment • Tilt angle range: 20±5 degrees (rear), 85±5degrees (front) • Rotate angle range: 90±5 degrees (left), 180±5 degrees (right) • Down/up: 150±20mm • Front/Rear: 350±20mm
2.2	Audio speakers		Stereo audio speakers
2.3	Monitor arm		Dual-wing floating monitor arm
2.4	Wheels		<ul style="list-style-type: none"> • Diameter: 125mm • 4 castors for total lock and break
2.5	Transducer port and holder		<ul style="list-style-type: none"> • Transducer ports with dust prevention: 5 active ports and 1 pencil probe port • Transducer holder: 7, 1 intracavitary transducer holder and 6 Transducer& ultrasound gel holders
2.6	Electrical power	3.3	<ul style="list-style-type: none"> • Voltage: 100-240VAC • Frequency: 50/60 Hz • Power consumption: 550 VA <p>Touch screen</p> <ul style="list-style-type: none"> • 15.6-inch high sensitivity anti-glare color touch screen • Resolution: 1920*1080 • Digital brightness and contrast adjustment through preset • Angle adjustable range: 50 degrees • Viewing angle: 178 degrees • Support touch screen gestures • Support either hand writing or with gloves on • Editable buttons: long press to add, delete or move buttons • Clinical scenario-based 3D/4D user interface • Digital TGC: 8 sliders • Short-cut switch of latest used probe & exams
2.7	Operating Environment		<ul style="list-style-type: none"> • Ambient temperature: 0-40°C • Relative humidity: 20%-85% (no condensation) • Atmospheric pressure: 700hPa-1060hPa
2.8	Storage & Transportation Environment		<ul style="list-style-type: none"> • Ambient temperature: -20-55 °C • Relative humidity: 20%-95% (no condensation) • Atmospheric pressure: 700hPa-1060hPa
2.9	System Noise		<p>≤26dB @25°C</p> <p>Touch gestures</p> <ul style="list-style-type: none"> • Swipe down/up: display/remove projected image on touch screen • Swipe horizontally: page up/down or review images/cine loops one by one • Swipe from left edge to right: display hidden menu on projected image. • Image parameter adjustment.
3	User Interface		
3.1	Control panel	3.4	<ul style="list-style-type: none"> • Brightness adjustable for the backlight of the control panel • Backlit QWERTY keyboard

- Measurement on projected image on touch screen
 - Zoom in/out the projected image on touch screen
 - Rotate or erase on projected 3D/4D image on touch screen
 - 8 user defined gestures using two fingers for more functions, such as freeze, print, activate specific imaging modes, measurements, and some other special functions.
- 3.5 System boot-up
- Boot-up from shut-down: <60 sec
 - Boot-up from stand-by: <15 sec
 - Shut-down: <30 sec
- 3.6 Comments
- Supports text input and arrow
 - Voice annotation: record voice as annotation for images and cine
 - Support freehand marking on touch screen
 - Adjustable text size and arrow size
 - Supports home position
 - Covers various application
 - User customizable
- 3.7 Bodymark
- More than 232 bodymarks for versatile application
 - User customizable
- ## 4 Imaging Parameters
- 4.1 B-mode
- Frame rate (max): 827f/s
 - A.Power
 - TGC: 8 sliders
 - Depth: 30 Levels, 1.5-40cm
 - Gain: 0-100, 1/step
 - FOV: On/Off
 - FOV Size
 - FOV Position: random adjustable
 - Image Quality:
 - SP5-1N, L13-3N: three levels of fundamental frequency, five levels of harmonic frequency
 - D7-2, SD8-1: four levels of fundamental frequency, five levels of harmonic frequency
 - L14-3W, C7-3: three levels of fundamental frequency, four levels of harmonic frequency
 - others: three levels of fundamental frequency, three levels of harmonic frequency
 - Persistence: 0-7, 1/step
 - Dyn Ra.:
 - ELC10-4C, DE11-3, SC5-1N, C5-1, C11-3: 30-160, 5/step; 160-180, 10/step; 180-260, 40/step
 - Others: 30-160, 5/step; 160-200, 10/step
 - Gray Map: 1-8, 1/step
 - Tint Map: Off, 1-8, 1/step
 - iClear: Off, 1-7, 1/step
 - iBeam: Off, 1-3, 1/step
 - Line Density: L, M, H, UH
 - L/R: On/Off
 - U/D: On/Off
 - Rotation: 0, 90, 180, 270
 - TSI: General, Muscle, Fat, Fluid
 - iTouch⁺: On/Off
 - LGC: 8 point
 - Dual Live: On/Off
 - Auto Merge: On/Off
 - H Scale: On/Off
 - Echo Boost: On/Off
 - Smooth: 0-6, 1/step
 - ExtImage: On/Off
 - ZoneVue (Phased array transducer cannot be adjusted)
 - Pan Zoom:
 - 0.8-1.2, 0.1/step
 - 1.2-2, 0.2/step
 - 2-4, 0.25/step
 - HD Scope: off, 1-3, 1/step
 - V1: 1: On/Off
 - Dehaze: 0~30
 - EdgeEnhance: 0~6, 1/step
 - HR Flow: On/Off
 - Single/Quad: On/Off
 - B Steer: 5 Levels
 - Trapezoid: On/Off
 - Ref Lines: On/Off
 - Free View: -45-45, 5/step

4.2	THI and PSH		
	<ul style="list-style-type: none"> • Patent PSH technology, obtains purer harmonic, better contrast resolution, higher SNR, exceptional high frequency harmonic • iClear available • Image quality: HPen/HGen/HRes or HPen/HPen-FFR/HGen/HRes/HRes-FFR (depends on transducers) • Echo Boost: On/Off 		<ul style="list-style-type: none"> • Color Map: V0-V10, VV0-VV9 • Flow State: L, M, H • Priority: 0%-100%, 1%/step • WF: 0-7, 7-1282 Hz • Line Density: L, M, H, UH • Dual Live: On/Off • Invert: On/Off • Auto Invert: On/Off • B/C Align: On/Off • velocity tag: On/Off • Packet Size: 0-3, 1/step • iTouch⁺: On/Off • Smart Track: On/Off • Glazing Flow: On/Off, L, M, H • Quad: On/Off • HR Flow: On/Off • ExtImage: On/Off
4.3	M-mode	4.5	Power Doppler Imaging
	<ul style="list-style-type: none"> • Gain: 0-100, 1/step • Speed: 25mm/s, 35mm/s, 50mm/s, 65mm/s, 100mm/s, 200mm/s • Dyn Ra.: <ul style="list-style-type: none"> - Others: 30-180, 5/step - ELC10-4C, DE11-3, SC5-1N, C5-1, C11-3: 30-160, 5/step; 160-180, 10/step; 180-260, 40/step • Gray Map: 1-8, 1/step • Tint Map: off, 1-8, 1/step • Display Format: V2: 3, V3: 2, H2: 3, V3: 1, FULL • M Soften: 0-4, 1/step • EdgeEnhance: 0-3, 1/step • Time Mark: On/Off • ZoneVue (Phased array transducer cannot be adjusted) 		<ul style="list-style-type: none"> • Velocity: max. 97.8cm/s; min. 1.0cm/s • PRF: 0.2-12.3kHz • Gain: 0-100, 2/step • scale • Steer • Quick Steer • ROI: random adjustable • Img Quality: <ul style="list-style-type: none"> - SC5-1N (Non-fetal heart), C5-1 (Non-fetal heart), L13-3N, SP5-1N (Cardiac), V11-3H, V11-3, D7-2, V11-3B, P8-2Ts: 5 levels - HR Flow: 1 level - Other: 3 levels • Persistence: 0-6, 1/step • Smooth: 0-6, 1/step • Color Map: P0-P3, dP0-dP3 • Flow State: L, M, H • Priority: 0%-100%, 1%/step • WF: 8 Levels • Line Density: L, M, H, UH • Dual Live: On/Off • Invert: On/Off • B/C Align: On/Off • Packet Size: 0-3, 1/step • iTouch⁺: On/Off • Smart Track: On/Off • Glazing Flow: On/Off, L, M, H • Quad: On/Off
4.4	Color Doppler Imaging		
	<ul style="list-style-type: none"> • Velocity: max. 97.8cm/s; min. 1.0cm/s • Max frame rate: 322f/s • PRF: 0.2-12.3kHz • Gain: 0-100, 2/step • Baseline: -8-8, 1/step • scale • Steer • Quick Steer • ROI: random adjustable • Img Quality: <ul style="list-style-type: none"> - SC5-1N (Non-fetal heart), C5-1 (Non-fetal heart), L13-3N, SP5-1N (Cardiac), V11-3H, V11-3, D7-2, V11-3B, P8-2Ts: 5 levels - HR Flow: 1 level - Other: 3 levels • Persistence: 0-6, 1/step • Smooth: 0-6, 1/step 		

		<ul style="list-style-type: none"> • HR Flow: On/Off • Dyn Ra.: 10-70, 5/step • ExtImage: On/Off
4.6	PW Mode	<ul style="list-style-type: none"> • PRF: 0.7-18.9kHz • Velocity: max. 729.2cm/s; min. 0.01cm/s • Gain: 0-100, 2/step • Baseline: -4-4, 1/step • scale • Steer • Quick Steer • Volume: 0%-100%, 2%/step • Angle: -89-89, 1/step • Quick Angle: -60, 0, 60 • SVD: random adjustable • Img Quality: <ul style="list-style-type: none"> - SC5-1N (Non-fetal heart), C5-1 (Non-fetal heart), L13-3N, SP5-1N (Cardiac), V11-3H, V11-3, D7-2, V11-3B, P8-2Ts: 5 levels - other: 3 levels • Speed: 25mm/s, 35mm/s, 50mm/s, 65mm/s, 100mm/s, 200mm/s • SV: 0.5-30mm • Dyn Ra.: 24-72, 2/step • Gray map: 1-10, 1/step • Tint Map: Off, 1-8, 1/step • Display format: V2: 3, V3: 2, H2: 3, V3: 1, FULL • Invert: On/Off • Auto Invert: On/Off • WF: 0-9, 14-1000Hz • Duplex/Triplex: On/Off • HPRF: On/Off • iTouch⁺: On/Off • T/F Res: 0-6, 1/step • Auto Calculate: On/Off • Auto Calc Cycle: 1-5, 1/step • Trace Sensitivity: 0-5, 1/step • Auto Calc Param • Trace Smooth: off, 1-4, 1/step • Trace Area: Above, Below, All • Auto Calc Loop
		4.8
4.7	CW Mode	<ul style="list-style-type: none"> • PRF: 0.2-100.0kHz • Velocity: max. 3850.0cm/s; min. 0.01cm/s
		<ul style="list-style-type: none"> • Gain: 0-100, 2/step • Baseline: -4-4, 1/step • Scale • Volume: 0%-100%, 2%/step • Angle: -89-89, 1/step • Quick Angle: -60, 0, 60 • Speed: 25mm/s, 35mm/s, 50mm/s, 65mm/s, 100mm/s, 200mm/s • Dyn Ra.: 24-72, 2/step • Gray map: 1-10, 1/step • Tint Map: Off, 1-8, 1/step • Display Format: V2: 3, V3: 2, H2: 3, V3: 1, FULL • Invert: On/Off • WF: 0-9, 5-27500Hz • T/F Res: 0-6, 1/step • SVD: random adjustable <p>Free Xros M/Free Xros CM</p> <ul style="list-style-type: none"> • Speed: 25mm/s, 35mm/s, 50mm/s, 65mm/s, 100mm/s, 200mm/s • Gray Map: 1-8, 1/step • Tint Map: off, 1-8, 1/step • Display format: V2: 3, V3: 2, H2: 3, V3: 1 • Angle • Delete (Free Xros CM) • Undo (Free Xros CM) • Edit (Free Xros CM) • Display: On/Off (Free Xros M) • show A: On/Off (Free Xros M) • show B: On/Off (Free Xros M) • show C: On/Off (Free Xros M) • ZoneVue (Phased array transducer cannot be adjusted) <p>Tissue Velocity/Energy Imaging (included in TDI option)</p> <ul style="list-style-type: none"> • Available on phased array • Max frame rate: 1510f/s • PRF: 0.4-13.9kHz • Velocity: max. 99.0cm/s; min. 4.9cm/s • Gain: 0-100, 2/step • Baseline: -8-8, 1/step • scale • ROI: random adjustable • Img Quality: 2 Levels • Persistence: 0-6, 1/step • Smooth: 0-6, /step • Dyn Ra.: 10-70, 5/step

	<ul style="list-style-type: none"> • Color Map: <ul style="list-style-type: none"> - TVI: TVV1-TVV10 - TEI: P0-P3, dP0-dP3 • Tissue State: L, M, H • Priority: 0%-100%, 1%/step • WF: 8 Levels • Line Density: L, M, H, UH • Dual Live: On/Off • Invert: On/Off • B/C Align: On/Off • velocity tag: On/Off • Packet Size: 0-3, 1/step 		<ul style="list-style-type: none"> • Priority: 0%-100%, 1%/step • Velocity tag: On/Off • Tissue state: L/M/H
4.10	Tissue Velocity Doppler (included in TDI option)	4.12	TDI QA
	<ul style="list-style-type: none"> • PRF: 0.7-18.9kHz • Velocity: max. 583.3cm/s; min. 0.01cm/s • Gain: 0-100, 2/step • Baseline: -4-4, 1/step • scale • Volume: 0%-100%, 2%/step • Angle: -89-89, 1/step • Quick Angle: -60, 0, 60 • SVD: random adjustable • Img Quality: 2 levels • Speed: 25mm/s, 35mm/s, 50mm/s, 65mm/s, 100mm/s, 200mm/s • Dyn Ra.: 24-72, 2/step • Gray map: 1-10, 1/step • Tint Map: Off, 1-8, 1/step • Display format: V2: 3, V3: 2, H2: 3, V3: 1, FULL • Invert: On/Off • WF: 10 Levels • Duplex/Triplex: On/Off • iTouch⁺: On/Off • T/F Res: 0-6, 1/step 	4.13	<ul style="list-style-type: none"> • Dedicated quantification tool for TDI velocity, strain, strain rate analysis • Ellipse ROI, Standard ROI • Up to 8 of ROI • Delete all • Delete current • ROI tracking: tracking ROI along with cardiac movement • Smooth: 1-7, 1/step • X scale: 1-5, 1/step • Std.Height: 1.5-50 mm • Std.Width: 1.5-50 mm • Std.Angle: -89-90 degrees • Export: export current data as CSV format file
4.11	Tissue Velocity Motion (included in TDI option)		Smart 3D
	<ul style="list-style-type: none"> • Display formats: V2:3, V3:2, V 3:1, H2:3, Full (V: vertical, H: horizontal) • Gain: 0-100 • M sweep speed: 6 steps • Smooth: 0-6, 1/ step • Color Map: 10 types • Image quality: 2 levels • Persistence: 0-6, 1/ step • Packet size: 0-3, 1/ step 		<ul style="list-style-type: none"> • Acquisition preparation <ul style="list-style-type: none"> - 3D/4D Scenario setting: Routine - Acquiring Methods: Rocked, Linear - Reset VOI - Flip VOI - Angle: 10°-80° - Distance: 10-200mm - Acquiring Time: 1s-20s • VR <ul style="list-style-type: none"> - 3D/4D Scenario: Routine-Surf., iLive Gen., iLive Transp., Skeleton - Reset: All, Ori, Curve - VOI: On, Off, Fixed - Active Quadrant: A, B, C, VR - VR orientation: 0°, 90°, 180°, 270° - Filp - Sync - 3D Ref - Orientation Assist - Threshold: 0%-100% - Opacity: 0%-100% - Brightness^{VR}: 0%-100% - Contrast^{VR}: 0%-100% - Smooth: 0-10 - Depth VR: Off, Black, Cyan, Blue, Rose - Hyaline^{iLive}: 0%-100%

- **Tint^{VR}**: Off, 1-8
 - **Degree**: 10°-80° ($\pm 20^\circ$)
 - **Distance**: 10-200mm ($\pm 20\text{mm}$)
 - **Zoom**: x0.2-x4.0
 - **MPR**
 - **Active Quadran**: A, B, C
 - **Gray Map**: 1-8
 - **Brightness^{MPR}**: 0%-100%
 - **Contrast^{MPR}**: 0%-100%
 - **iClear**: Off, 1-7
 - **Zoom**: x0.2-x4.0
 - **Tint^{MPR}**: Off, 1-8
 - **Thickness**: 0mm-30mm
 - **Adv.**
 - **Direction**: Up/Down, Left/Right, Font/Back, Down/Up, Right/Left, Back/Front
 - **VR Refine**: Off, 1-7
 - **Surf.ENH**: 0-7
 - **MgaiClean**: Off, Low, Mid, Hight, Max
 - **Shading**: 0-10
 - **Grad View**
 - **Hyaline**
 - **Hyaline^{iLive}**: 0%-100%
 - **Inversion**
 - **A3:1**
 - **Move Light**
 - **Distance**: 10-200mm ($\pm 20\text{mm}$)
 - **Threshold**: 0%-100%
 - **Brightness^{VR}**: 0%-100%
 - **Depth VR**: Off, Black, Cyan, Blue, Rose
 - **Smooth**: 0-10
 - **Opacity**: 0%-100%
 - **Contrast^{VR}**: 0%-100%
 - **Tint^{VR}**: Off, 1-8
 - **Degree**: 10°-80° ($\pm 20^\circ$)
 - **Main Render**: Surface, Max, Min, X Ray, iLive
 - **Sub Render**: Surface, Max, Min, X Ray
- Spine, Brain, Long Bone, Face, Endometrium, Pelvic**
- Routine**
- iLive Pro**
- Bone**
- Tissue**
- Refresh**
- Reset VOI**
- Flip VOI**
- Angle**: 10°-90°
- Quality**: low1, low2, mid, high1, high2
- STIC**
- VR:**
- **Reset**: All, Ori, Curve
 - **VOI**: On, Off, Fixed
 - **Active Quadrant**: A, B, C, VR
 - **VR orientation**: 0°, 90°, 180°, 270°
 - **Filp**
 - **Sync**
 - **Smart Face**
 - **3D Ref**
 - **Orientation Assist**
 - **Threshold**: 0%-100%
 - **Opacity**: 0%-100%
 - **Brightness^{VR}**: 0%-100%
 - **Contrast^{VR}**: 0%-100%
 - **3D iClear^{VR}**: Off, 1-7
 - **Smooth**: 0-10
 - **Depth VR**: Off, Black, Cyan, Blue, Rose
 - **Tint^{VR}**: Off, 1-8
 - **Hyaline^{iLive}**: 0%-100%
 - **Zoom**: x0.2-x4.0
 - **Face+**: Off, 1-3
 - **FaceContact**: -15 - 15
 - **Auto Play**: Stop, x1/3, x1/2, x1, x2, x3
 - **Frame**: 1-N/N (N indicates the number of all frames)
 - **3D/4D Scenario**: Smart Scene 4D-Surf., Skeleton, iLive Gen., iLive Bone

4.14 4D

- Available on all volume transducers
 - Static 3D and real time 4D
- Acquisition preparation:**
- 4D frame rate: max. 80 vps
 - Smart Scene3D:
- Routine-Surf., Surf.Enh., iLive Gen., Skeleton**
- iLive Pro-iLive Transp., Spot Light, Inner Light, Multi-Light**

- Bone-Skeleton, Skeleton.Surf, iLive
- Bone, Skeleton.Smooth
- Tissue-Surf.Detail, Surf.Smooth,
- Soft.Inversion, iLive Tissue
- MPR:
 - Active Quadrant: A, B, C
 - Gray Map: 1-8
 - Brightness^{MPR}: 0%-100%
 - Contrast^{MPR}: 0%-100%
 - 3D iClear^{MPR}: Off, 1-7
 - Tint^{MPR}: Off, 1-8
 - Zoom: x0.2-x4.0
 - Thickness: 0mm-30mm
 - Auto Play: Stop, x1/3, x1/2, x1, x2, x3
 - Frame: 1-N/N (N indicates the number of all frames)
- Adv.:
 - Direction: Up/Down, Left/Right, Font/Back, Down/Up, Right/Left, Back/Front
 - VR Refine: Off, 1-7
 - Surf.ENH: 0-7
 - MgaiClean: Off, Low, Mid, Hight, Max
 - Inversion
 - Move Light
 - Main Render: Surface, Max, Min, X Ray, iLive
 - Sub Render: Surface, Max, Min, X Ray
 - Shading: 0-10
 - Hyaline
 - Hyaline^{iLive}: 0%-100%
 - Grad View
 - Mix: 100%/0%-0%/100%
 - Threshold: 0%-100%
 - Opacity: 0%-100%
 - Brightness^{VR}: 0%-100%
 - Contrast^{VR}: 0%-100%
 - 3D iClear: Off, 1-7
 - Smooth: 0-10
 - Depth VR: OFF, Black, Cyan, Blue, Rose
 - Tint^{VR}: Off, 1-8
 - Auto Play: Stop, x1/3, x1/2, x1, x2, x3
- Frame: 1-N/N (N indicates the number of all frames)
- Tools
 - Edit:
 - Rubber
 - Polygon
 - Line
 - Contour
 - Rect
 - Undo
 - Undo All
 - Eraser Diam.: 8-80
 - Auto Play: Stop, x1/3, x1/2, x1, x2, x3, Static 3D and real time 4D support
 - Frame: 1-N/N (N indicates the number of all frames), Static 3D and real time 4D support
- 3D Layout:
 - Niche: Inner, Outer
 - 3Slice
 - Active Quadrant: A, B, C, Niche/ 3Slice
 - Auto Play: Stop, x1/3, x1/2, x1, x2, x3, Static 3D and real time 4D support
 - Frame: 1-N/N (N indicates the number of all frames), Static 3D and real time 4D support
- Auto rotation:
 - Save AVI to USB
 - Set Start
 - Set End
 - Direction: Left/Right, Up/Down
 - Step: 1°-15°
 - Speed: 1-6
 - Quick Angle: 30°-180°
- 3D Print
 - Quality: Low, Mid, High
 - Generate Mesh
 - File Format: .stl, .obj, .ply, .3mf, .off
 - Save mesh to USB
 - Threshold: 0%-100%
 - 3D iClear^{VR}: Off, 1-7, Static 3D and real time 4D support
 - Smooth: 0-10
- 3D Reference Point
 - Input
 - Display: Point, H Line, V Line

- Delete All
 - Hide All
 - Return
 - Distance: 10°-80° ($\pm 20^\circ$), Smart 3D support
 - Degree: 10mm-200mm ($\pm 20\text{mm}$), Smart 3D support
 - Auto Play: Stop, x1/3, x1/2, x1, x2, x3, Static 3D and real time 4D support
 - Frame: 1-N/N (N indicates the number of all frames), Static 3D and real time 4D support
 - Color 3D
 - Available on volume transducers
 - Supports Color and Power mode
 - STIC
 - Acquiring Time: 7.5s-17.5s
 - Angle: 10°-85°
 - Reset VOI
 - Flip VOI
 - Estimated Fetal Heart Rate:
ACCEPT HR
REJECT HR
REJECT ALL
 - iPage⁺
 - Reset Ori
 - RefMPR: A Plane, B Plane, C Plane
 - Sync MPR: A, B, C
 - Quick Rotation: 90°, -90°, -180°
 - Slice Layout: 2*2, 3*3, 4*4, 5*5
 - Tint: Off, 1-8
 - Brightness: 0%-100%
 - Contrast: 0%-100%
 - Thickness: 0mm-30mm
 - 3D iClear: Off, 1-7
 - Slices Number: 3-25
 - Spacing: 0.5mm-10.0mm
 - Line(Start)
 - Line(End)
 - Auto Play: Stop, x1/3, x1/2, x1, x2, x3
 - Frame: 1-N/N (N indicates the number of all frames)
 - SCV⁺
 - SCV⁺
 - CMPR
 - Reset: Reset All
 - Orientation Assist
 - Active Quadrant: A, B, C
 - 3D iClear: Off, 1-7
 - Tint: Off, 1-8
 - Brightness: 0%-100%
 - Contrast: 0%-100%
 - Thickness: 0mm-30mm
 - Opacity: 0%-100%
 - Main: Surface, Max, Min, X Ray
 - Sub: Surface, Max, Min, X Ray
 - Mix: 0/100%-100/0%
 - Auto Play: Stop, x1/3, x1/2, x1, x2, x3
 - Frame: 1-N/N (N indicates the number of all frames)
- CMPR**
- Reset: Undo Last, Reset Curve
 - CMPR
 - Flip: Thickness, Direction
 - Active Quadrant: Ref. Image (A, B, C), 1, 2, 3
 - Trace Options: Line, Trace, Spline
 - Quick Rotation: 0°, 90°, 180°, 270°
 - Hide Annotation
 - Tint: Off, 1-8
 - Brightness: 0%-100%
 - Contrast: 0%-100%
 - Thickness: 0mm-30mm
 - Opacity: 0%-100%
 - 3D iClear: Off, 1-7
 - Line Extension
 - Rotate RL
 - Auto Play: Stop, x1/3, x1/2, x1, x2, x3
 - Frame: 1-N/N (N indicates the number of all frames)
- iLive**
- Classic
 - Int Point
 - Ext Point
 - User1
 - Paralled
 - Torch
 - 3-light
 - User2
 - Light
 - Copy to: User1, User2
 - Return

- Move Light
- VL Dist.^{iLive}: 0.0-5.0
- VL Angle^{iLive}: 0%-100%
- Brightness^{VR}: 0%-100%
- Contrast^{VR}: 0%-100%
- Depth VR: Off, Black, Cyan, Blue, Rose
- VL Bright.^{iLive}: 0%-100%
- Tint: 0-4
- Hyaline^{iLive}: 0%-100%
- VL Hue^{iLive}: 0%-100%
- VL Sat.^{iLive}: 0%-100%
- Degree: 10°-80° ($\pm 20^\circ$), Smart 3D support
- Distance: 10mm-200mm ($\pm 20\text{mm}$), Smart 3D support
- Smart Planes CNS
 - Auto Comment
 - Reset: All Planes, Current Planes
 - Measure: Auto Measure, Edit (Off, BPD, OFD, HC, TCD, CM, LVW)
 - Accept Result
 - Accept
 - Cancel
 - Smart Planes: TCP, TTP, MSP, TVP
 - MSP Adjust: A, B, C
 - Hide Measure
 - Hide Lines
 - Font Size: Small, Medium, Large
 - Tint: Off, 1-8
 - Brightness: 0%-100%
 - Contrast: 0%-100%
 - Thickness: 0mm-30mm
 - 3D iClear: Off, 1-7
- Smart Face
 - Recognize fetal face automatically and then display the face in a recommended viewing angle
 - AutoDirect
 - MixRender: 0-3
 - EraseSize: 0-2
 - For-Rubber
 - Re-Rubber
 - SubTint: 1-4
 - Face+: Off, 1-3
 - FaceContact: -15-15
- Smart FLC (Smart Follicle)
 - Automatic follicle calculation in 2D/3D
 - Calc
 - Edit ROI
 - Active Quadrant: A, B, C, Follicle, Left, Right
 - Edit: Edit On, Edit Off, Divide, Merge, Add/Del
 - Undo: Undo, Redo, Undo All
 - Save to Report
- Smart Scene 3D
 - As an intelligent scenario-oriented volume scan technique, it automatically enables the identification of tissue characteristics and provides an organ-specific diagnosis with full-stack intelligence from imaging optimization, to planes acquisition, quantification and with automated workflow throughout whole procedure.
 - Available in fetal brain, fetal face, spine, long-bone, endometrium, and pelvic floor.
 - Not available on Smart 3D mode
 - Adjust ROI size and position automatically and activate appropriate render modes according to the recognized anatomical structure of certain organs
- Smart Volume
 - Smart-V ROI
 - Edit ROI
 - Calc
 - Active Quadrant: A, B, C, Smart-V
 - ROI Range: 0-2
 - Tint: Off, 1-8
 - Brightness: 0%-100%
 - Contrast: 0%-100%
 - Thickness: 0mm-30mm
 - Trace Mode: Smart Trace, Spline, Trace
 - Display: 2D, 2D&C
 - Edit: Off, On (Add, Delete, ReCalc)
 - Analysis: Shell-Off, Inside, Outside, Sym.
 - Accept Result

- Smart-V Reset Ori	4.15	Smart Track
- shell thickness: 0-20mm		<ul style="list-style-type: none"> Available on linear probes
- Undo: Undo, Redo, Undo All		<ul style="list-style-type: none"> Enable the function under Color/Power mode, the angle and the position of the ROI are adjusted automatically.
Smart-V Trace		<ul style="list-style-type: none"> Enable the function under Color/Power+PW mode, the angle and the position of the PW sampling line, SV size, SV angle and SV position are adjusted automatically
- Trace On MPRs		RIMT (RF-Data based IMT)
- Calc		<ul style="list-style-type: none"> Side: left/right
- Active Quadrant: A, B, C, Smart-V		<ul style="list-style-type: none"> Calculation of 6 RIMT values, RIMT average value, SD and ROI W
- Tint: Off, 1-8		
- Brightness: 0%-100%		
- Contrast: 0%-100%		
- Thickness: 0mm-30mm	4.16	iScape View
- ROI Range: 0-2		<ul style="list-style-type: none"> Actual size
- Trace Mode: Smart Trace, Spline, Trace, Control Point		<ul style="list-style-type: none"> Fit size
- Display: 2D, 2D&C		<ul style="list-style-type: none"> Ruler: On/Off
- Edit: Off, On (Add, Delete, ReCalc)	4.17	<ul style="list-style-type: none"> Tint map: Off; 8 types
- ReCalc		<ul style="list-style-type: none"> Rotation: 0~355 degrees, 5/step
- Analysis: Shell-Off, Inside, Outside, Sym.		
- Accept Result		
- Smart-V Reset Ori		
- shell thickness: 0-30mm		
- Undo: Undo, Redo, Undo All	4.18	iNeedle
Smart V Vocal		<ul style="list-style-type: none"> B/iNeedle: On/Off
- 1.Reference Line		<ul style="list-style-type: none"> Needle Dir.: Auto/Left/Right
- 2.Trace OnSlice		<ul style="list-style-type: none"> LineDensity: L, M, H, UH
- 3.Cal Vol		
- Active Quadrant: A, B, C, Smart-V	4.19	UWN+ Contrast Imaging
- Tint: Off, 1-8		<ul style="list-style-type: none"> Available on C5-1, SC5-1N, DE11-3, V11-3, V11-3B, V11-3H, V11-3HB, L9-3, L13-3N, L14-3W, L16-4Hs, L20-5s, SP5-1N, and C7-3 probes
- Brightness: 0%-100%		<ul style="list-style-type: none"> Gain: 0-100, 1/step
- Contrast: 0%-100%		<ul style="list-style-type: none"> FOV
- Thickness: 0mm-30mm		<ul style="list-style-type: none"> FOV Position: random adjustable
- Display: 2D, 2D&C		<ul style="list-style-type: none"> Image Quality: 3 Levels
- Trace Mode: Smart Trace, Spline, Trace, Control Point		<ul style="list-style-type: none"> Persistence: 0-7, 1/step
- Slice Num: 6-30		<ul style="list-style-type: none"> Gray Map: 1-8, 1/step
- Edit: On, Off		<ul style="list-style-type: none"> Tint Map: Off, 1-8, 1/step
- Win Format: 1*1, 3*2, 3*3, 4*3		<ul style="list-style-type: none"> iClear: Off, 1-7, 1/step
- Update Vol		<ul style="list-style-type: none"> Line Density: L, M, H, UH
- Analysis: Shell-Off, Inside, Outside, Sym.		<ul style="list-style-type: none"> L/R Flip: On/Off
- Accept Result		<ul style="list-style-type: none"> U/D Flip: On/Off
- Smart-V Reset Ori		<ul style="list-style-type: none"> Rotation: 0, 90, 180, 270
- Shell thickness: 0-30mm		<ul style="list-style-type: none"> LGC: 8 point
- Undo: Undo, Redo, Undo All		<ul style="list-style-type: none"> Dual Live: On/Off
- Previous Slice		<ul style="list-style-type: none"> Smooth: 0-6, 1/step
- Next Slice		<ul style="list-style-type: none"> Mix: <ul style="list-style-type: none"> - Dual Live on: Contrast/C&T

	<ul style="list-style-type: none"> - Dual Live off: Contras/C&T/Tissue • Timer1: On/Off • Timer2: On/Off • Destruct: On/Off • Destruct Time: <ul style="list-style-type: none"> - Phased array transducer: 200-2000, 90/step - Curved array transducer, Linear array transducer: 500-2000, 75/step • DestructAP: -30-0, 0.3/step • Max Frame Rate • MFE: On/Off • MFE Period: 0.1s, 0.2s, 0.4s, 0.6s, 0.8s, 1.0s, MAX • Restro Capture: On/Off • Pro Capture: On/Off • CEUSPos: On/Off • MarkLine: On/Off • iTouch⁺: On/Off 	4.23	Volume CEUS <ul style="list-style-type: none"> • Display: Contrast, Tissue • Reset: All, Ori, Curve • VOL: On, Off, Fixed • Active Quadrant: A, B, C, VR • Rotation: 0°, 90°, 180°, 270° • Flip • Sync • Threshold: 0%-100% • Opacity: 0%-100% • Brightness: 0%-100% • Contrast: 0%-100% • Smooth: 0-10 • Main: Surface, Max, Min, X-ray, iLive • Sub: Surface, Max, Min, X-ray, iLive • Mix: 100%/0%-0%/100% • Timer1 • Timer2 • Set Begin • Set End • Jump To First • Jump To Last • Auto Play: Stop, ×1, ×2, ×3, ×1/2, ×1/4 • Frame • Rubber: 20, 40, 60, 80 • Line • Contour
4.20	Contrast Imaging QA		
	<ul style="list-style-type: none"> • Support Time-Intensity Curve analysis • Table display: display data in table • Up to 8 ROIs • Delete all • Delete current • Copy ROI • Fit curve • Raw curve • Motion tracking: Reduce the effect of tissue movement • X scale: 1-5, 1/step • Auto play • ROI Type: Trace ROI, Ellipse ROI • Export: export current data as CSV format file 	4.24	Strain Elastography <ul style="list-style-type: none"> • Available on V11-3H, V11-3, V11-3B, V11-3HB, L13-3N, L14-3W, L9-3, L16-4Hs, L20-5s, and DE11-3 probes • Opacity: 0-5, 1/step • Map: E1-E6 • Smooth: 0-5, 1/step • ROI: random adjustable • Invert: On/Off • Depth • Display Format: V1: 1, H1: 1, FULL • Strain Scale: 0~5, 1/step • Map Position: 0%~100%, 5%/step • Dyn Ra.: 0~5, 1/step • Strain Mode: 0~1, 1/step • E Sensitivity: 0~5, 1/step
4.21	LVO		
	<ul style="list-style-type: none"> • Only available in LVO exam mode • Dedicated left ventricle contrast imaging tool 		
4.22	VLM Contrast		
	<ul style="list-style-type: none"> • Only available in cardiac exam mode • Enhances echo reflection by using contrast agent to perform myocardial analysis based on echocardiography technique 	4.25	STE Imaging (Sound Touch Elastography Imaging) <ul style="list-style-type: none"> • Available on C5-1, SC5-1N, L14-3W, L13-3N, L20-5s, and L9-3 probes

	<ul style="list-style-type: none"> Display Format: V1: 1, H1: 1, FULL RLB Map: On/Off, RLB, RLB&E, RLB&B&E iLayering: On/Off Invert: On/Off HQElasto: On/Off E Quality: Pen, Gen, Res Elas.Metric: E, Cs, G Scale: 30 levels Opacity: 0-5, 1/step Map: E1-E3, 1/step ROI: random adjustable Filter: 0, 1 RLB View: On/Off Map Position: 0%~100%, 5%/step Smooth: 0~2, 1/step Persistence: 0~2, 1/step M-STB Index: On/Off M-STB Sensi.: 0~4, 1/step E Bar: Mean, Max, Min, SD E Avg: 8step Fixed ROI Save All: On/Off Lesion: off, 1~10 		
4.26	STQ Imaging (Sound Touch Quantification Imaging)	4.28	<ul style="list-style-type: none"> Adjustment for the border of endocardium by single point or multi points Adjust Frame Layout: Dual/ Single Diastole FR Systole FR Volume curve: On/Off
4.27	AutoEF		<p>TTQA</p> <ul style="list-style-type: none"> Tissue tracking quantitative analysis Available on SP5-1N and P8-2 probes Mandatory ECG connection before TTQA cine acquisition Six views for analysis: ALAX, A4C, A2C, PSAXB, PSAXM, PSAXAP Reload: reload cine again for new study Edit: modify trace points Start tracking Accept & compute: start tracking myocardium movement when user accept trace result Display effect: 0/1; at 1, tracking in velocity vector arrow; at 0, tracking in dots Trace method: 3 point or manual for ALAX, A4C, A2C; manual for PSAXB, PSAXM, PSAXAP Bull's eye: trace result in bull's eye model Valve's open and close time index: MVC, MVC; AVC, AVO, MVO Data export: export data in CSV file Cycle: ECG triggered cardiac cycle recognition for analysis; adjustable Auto play: stop, X1/10, X1/5, X1/4, X1/3, X1/2, X1, X2, X3 Thickness: 1-30mm, 1mm/step; adjust trace thickness Track point: 20-40, 1/step Parameter: Volume, Speed, Displacement, L Strain, L Strain R, T Strain, T Strain R, Area, R Strain, R Strain R, C Strain, C Strain R, C Rotation, C Rotation R Smooth: 0-4, 1/step Trace method: 3 point, manual Tracking cycles: 1, 3 Select Cycle: select among 3 cycles when the Tracking Cycle is set to 3

	<ul style="list-style-type: none"> Average Cycle: On/Off Display style: All, Endo., Myo., Epi., Curve Display Torsion & Torsion Rate Curve LGC adjustment Data Export: export current data as CSV format file 	4.31	<ul style="list-style-type: none"> Position: 0-100% Curve Disp: Adjust the height of the curve
4.29	Stress Echo		<p>Smart Pelvic</p> <ul style="list-style-type: none"> VR <ul style="list-style-type: none"> Auto Measure Accept Result Refresh Oir Set Lables: Rest, Contraction, Valsalva Active Quadrant: A, B, C, VR Tools: Edit, Hide, Undo, Undo All VR Refine: Off, 1-7 VOI Thickness: 10mm-30mm Threshold: 0%-100% Opacity: 0%-100% Brightness: 0%-100% Contrast: 0%-100% 3D iClear: Off, 1-7 Timer: On, Off Meas Parm Smooth: 0-10 Auto Play: Stop, x1/3, x1/2, x1, x2, x3 Frame: 1-N/N (N indicates the number of all frames) Multi-plane Mode Meas Parameters: <ul style="list-style-type: none"> Rest: LH Area, LH Circ, LH AP Diam, LH Lateral Diam, LUG Contraction: LH Area, LH Circ, LH AP Diam, LH Lateral Diam, LUG Valsalva: LH Area, LH Circ, LH AP Diam, LH Lateral Diam, LUG CMPR <ul style="list-style-type: none"> Set Ref. Lines Accept Result Auto Measure Refresh Ori Set Lables: Rest, Contraction, Valsalva Active Quadrant: Ref. Image (A, B, C), Open, Closed, Closing Tools: Edit, Hide, Undo, Undo All 3*3 3D iClear: Off, 1-7 Line Extension Rotate RL
4.30	R-VQS		<ul style="list-style-type: none"> Side: Left, Right Speed: 6 levels

	<ul style="list-style-type: none"> - Opacit: 0%-100% - Brightness: 0%-100% - Contrast: 0%-100% - Tint: Off, 1-8 - Thickness: 0mm-30mm - Spacing: 0.5mm-10.0mm - Auto Play: Stop, x1/3, x1/2, x1, x2, x3 - Frame: 1-N/N (N indicates the number of all frames) 	4.36	Smart ABDvue
4.32	Smart Fetal HR (Fetal Heart Rate)		<ul style="list-style-type: none"> • Measure the fetal heart rate automatically on B and M modes • Smart FHR cycle: 1-3 • Alignment Out-Plane
4.33	Smart Thyroid	4.37	Smart Echovue
	<ul style="list-style-type: none"> • Side: Left, Right, Isthmus • Select nodule • Add nodule/Delete nodule • Detect view • Add/delete malignant signs view • Display/Hide measurement results • Analyze Nodule • Confirm Nodule • Review Nodule Analysis 		<ul style="list-style-type: none"> • Automatic abdomen plane recognition and scanning guidance. Smart ABDvue automatically recognizes the standard abdomen view, captures images/clips, helping standardize the quality of examination. • C5-1and SC5-1N in abdomen exam mode support this function. • Abdomen exam mode standard view: Portal Vein, Sag Liver-Rt Kidney,1st Porta Hepatis, Sag Lt Lobe-Caudate, Sag Lt Lobe-AAo, Sag LHV-IVC, Long GB, Lt Portal Vein,LA Pancreas, 2nd Porta Hepatis.
4.34	Smart Breast	4.38	iBeam
	<ul style="list-style-type: none"> • Side: Left, Right • Select Mass • Add Mass/Delete Mass • Detect view • Add/delete malignant signs view • Display/Hide measurement results • Analyze Mass • Confirm Mass • Review Mass Analysis 		<ul style="list-style-type: none"> • Spatial compound imaging • Off, 1-3, 1/step
4.35	iScanHelper	4.39	iClear
	<ul style="list-style-type: none"> • Tutorial function as a guidance to show basic scanning skill with graphic of probe position, schematic of anatomy and example clinical image • Support ABD, SMP, URO, OB, GYN applications • Support broadcasting the scanning skill in multi languages 		<ul style="list-style-type: none"> • Speckle suppression imaging • Available on B, 3D, 4D mode
		4.40	iTouch⁺
			<ul style="list-style-type: none"> • B-mode: Gain, TGC, LGC, Dehaze • Color: Gain • Power: Gain • PW: Scale, PRF • Contrast imaging: Gain, TGC, LGC
		4.41	Echo Boost
			<ul style="list-style-type: none"> • Available in cardiac exam mode when using a phased array probe • Improve the homogeneity through the whole field of view • Better noise control in cardiac chambers and muscles

4.42	Zoom		<ul style="list-style-type: none"> • Zoom: <ul style="list-style-type: none"> - Spot Zoom - Res Zoom - Pan zoom: 0.8x-4x • iZoom: convertible 3 steps; normal image, zoom standard area, zoom only image area 	<ul style="list-style-type: none"> • Maximum 4D cine memory: 12838 volumes (D7-2) • Retrospective storage (1-120s pre-settable) and prospective storage (1-480s pre-settable) • Frame compare: displays one cine in dual format and allows frame by frame compare side by side • Cine compare: compare cines which are saved in same imaging mode • Jump to first and jump to last: one keystroke go to first or last frame in the cine
4.43	QSave		<ul style="list-style-type: none"> • Quickly save image parameter setting after image adjustment done • Support Save, Create, Restore • 3D/4D Preset Manager <ul style="list-style-type: none"> - The scenarios and subpresets can be renamed, restored, deleted, added, set to active, or moved - Show scenario and subpreset parameters - Provide multiple groups of preset 3D/4D parameters based on different application scenarios to quickly obtain expected image effect 	5.2
4.44	iCompare			Raw data processing
4.45	DVR			<ul style="list-style-type: none"> • B-mode TGC Gain Dyn Ra. Gray Map Tint Map iClear L/R U/D Rotation LGC Dual Live Auto Merge H Scale Echo Boost Smooth ExtImage Zoom V1:1 Dehaze Edge Enhance Single/Quad Ref Lines
5	Cine Review and Raw Data Processing			<ul style="list-style-type: none"> • M-mode: <ul style="list-style-type: none"> Gain Speed Dynamic Range Gray Map Tint Map Display format Edge Enhance Time Flag • Color: <ul style="list-style-type: none"> Gain
5.1	Cine review		<ul style="list-style-type: none"> • Available in all modes • Frame by frame manual cine loop review or auto playback with variable speed • Maximum cine memory up to 32577 frames (B storage server) or 163.84s (M storage server) (depends on the mode) 	

Baseline			
Smooth			
Color Map	6.3	Smart OB	<ul style="list-style-type: none"> Data of IVF history exams can be checked in the IVF report.
Priority			
Dual Live			
Invert			
velocity tag			
Quad			
ExtImage			
• PW:			
Gain	6.4	Smart NT	
Baseline			
Volume			
Angle			
Quick Angle	6.5	Smart HIP	
Dynamic range			
Gray map			HIP auto measurement
Tint Map	6.6	Smart Trace	
Display format			Measures the lengths of major axis and minor axis, area and circumference of a closed region on the image semi- automatically
Invert			
WF			
T/F Res			
Auto Calculate	6.7	CPP	
Auto Calc Cycle			
Trace Sensitivity			
Auto Calc Parameter	6.8	Smart HRI	Measures the pixel proportion of blood flow signal in the region of interest
Trace Smooth			
Trace Area			
Auto Calc Loop	6.9	Smart Bladder	
			Auto measure three diameters and calculate the bladder volume
6 Measurement/Analysis and Report			
NOTICE:			
For general measurement, automatic measurement, and clinical measurement packages, see the Appendix.	6.10	Smart Calc	
			Automatic trace, measurement and calculation tool
	6.11	Report	
			<ul style="list-style-type: none"> Specific report template by application Editable value in report Images selectable Anatomy information for vascular and OB report Editing through iReport <ul style="list-style-type: none"> User-defined report template Selecting report modules Adding/removing measurement items from the report Changing report layout
6.1 IMT			
• Intima-Media Thickness Measurement			
• Automatic detection of IMT when ROI is set			
• Support CCA, ICA, ECA, Bulb IMT			
• Near wall and far wall detection			
• Angle selectable			
6.2 IVF			
• The uterus and follicle growth curve can be displayed in the IVF report.			

	<ul style="list-style-type: none"> • Load/save comment • Viewing history reports • Preview and printing reports • Able to Export as PDF/RTF file • V-Mapping Add anatomical graphics for illustration • Mini report <ul style="list-style-type: none"> - Quickly displaying Mini report in the thumbnail area of the main screen - Including both general measurement and application measurement results - Support deleting measurement results 	<ul style="list-style-type: none"> - DICOM Urology SR - DICOM Pediatric SR - DICOM Small Parts SR - DICOM Abdomen SR - TDI QA - Contrast Imaging QA - Tissue Tracking QA - Niche - iLive - iPage⁺ - IVF - AutoEF - Smart Planes CNS 	
6.12	iWorks	<p>Not all measurements are listed in this part; For more detailed information please refer to User Manual</p>	
	<ul style="list-style-type: none"> • Auto workflow protocol • Templates are user configurable • Functions: pause, stop, replace, repeat, skip, insert single step, return and continue, steps in thumbnail, iNSert another template • iWorks setup mode: B; B/B (Dual Live); Dual B/B; Color; B/Color (Dual Live); Power; B/Power (Dual Live); B + PW; Color + PW; Power + PW; B + CW; Color + CW; B+M; B+TVI; TVI+TVD; iScape View. • iWorks setup annotation: support up to 2 annotations, location and font size are configurable. • iWorks setup bodymark: select existing library, and transducer indicator is pre-settable • iWorks setup measurement: select existing measurement library • Template import and export are available • Support create user-defined iWorks protocol 	<p>7</p> <p>7.1</p> <p>7.2</p>	<h3>Exam Storage and Management</h3> <p>Exam storage</p> <ul style="list-style-type: none"> • 1TB HHD (Hard Disk Drive) Capable to store up to approximate 2297011 single frames • 512G SSD (Solid State Drive) Capable to store up to approximate 982869 single frames • 1TB SSD (Solid State Drive) Capable to store up to approximate 2297011 single frames • Direct digital storage of single frame and cine 2D, color and Doppler <p>Exam management</p> <ul style="list-style-type: none"> • iStation workstation dedicated for patient exam management • Patient exam query/retrieve • Support review of current and past exam • New exam, Activate exam, Continue exam functions, End exam are available • Support measurements and calculations on archived exam and images • Export images in BMP/JPG/TIFF/DCM/AVI/MP4 format • Support backup/send to USB devices, DVD-RW, CD-R, DVD+R, DVD-R, DVD+RW media
6.13	UltraView		
	<ul style="list-style-type: none"> • Components: <ul style="list-style-type: none"> - DICOM Basic - DICOM Query/Retrieve - DICOM OB/GYN SR - DICOM Cardiac SR - DICOM Vascular SR - DICOM Breast SR 		

8	Connectivity		
8.1	Ethernet Network Connection		<ul style="list-style-type: none"> Cable connection Wireless connection: built-in wireless adapter
8.2	DICOM 3.0		<ul style="list-style-type: none"> DICOM basic <ul style="list-style-type: none"> Verify (SCU, SCP) Task management DICOM storage DICOM print DICOM storage commitment DICOM media storage (including DICOM DIR) DICOM Worklist DICOM Query/Retrieve DICOM Modality Performed Procedure Step - MPPS DICOM OB/GYN structure report DICOM Cardiac structure report DICOM Vascular structure report DICOM Breast structure report DICOM Abdomen structure report DICOM Small Parts structure report
8.3	iStorage (included in UltraAssist)	8.6	<p>Direct network storage tool between ultrasound system and personal computer</p>
8.4	MedSight	9	<ul style="list-style-type: none"> An interactive app that lets you transfer clinical images straight from Mindray Ultrasound system to a smart device, such as mobile phone or tablet PC Needs to be installed on mobile terminal Transfer images or clips from system to mobile terminal through Wi-Fi Support both iOS (7.0 and above) and Android (4.0 and above) system For iOS powered smart device: DICOM is mandatory For Android powered smart device: DICOM not necessary
8.5	MedTouch	9.1	<ul style="list-style-type: none"> Connect Ultrasound machine to smart devices based on Android and iOS
			<p>system, such as tablet PC or mobile phone. Remote control of Ultrasound machine, review of patient information, and tutorial software iScanHelper study on smart devices</p> <ul style="list-style-type: none"> Support Android and iOS powered smart devices <ul style="list-style-type: none"> Android 4.0 and above iOS 7.0 and above DICOM not necessary <p>u-Link</p> <p>The u-Link can be used to connect to software that supports the u-Link protocol (such as the MiCO+ Remote Imaging System).</p>
			<h3>Transducers</h3> <p>Curved array</p> <ul style="list-style-type: none"> C5-1 <ul style="list-style-type: none"> Application: Abdominal, Gynecology, Obstetrics, Vascular, Nerve, Musculo-skeletal, Urology, Thoracic/Pleural, Small Organ Advanced Function: iScape View, Free Xros M, Color M, Smart 3D, Contrast Imaging (Abdomen), STE&STQ (Abdomen) Bandwidth: 1.2-6.0 MHz Number of Elements: 128 FOV (max): 74° Convex Radius: 60 mm Depth: 4.0-40.0 cm Physical Footprint: 76.5 mm*28 mm Footprint: 64.9 mm*16.2 mm B-mode Frequencies: 1.2-3.8, 1.7-5.2, 2.0-6.0 MHz Harmonic Frequencies: 4, 5, 6 MHz Color Frequencies: 1.8, 2.0, 2.5, 3.0, 3.2 MHz HR Flow: 3.5 MHz PW Frequencies: 1.8, 2.0, 2.5, 3.0, 3.2 MHz Biopsy Guide: NGB-022, multi angle, reusable, LPUBKG60 (disposable) C11-3

- Application: Abdominal, Cephalic, Cardiac, Vascular, Pediatric
 - Advanced Function: iScape View, Free Xros M, Color M, Smart 3D
 - Bandwidth: 2.6-12.8 MHz
 - Number of Elements: 128
 - FOV (max): 114°
 - Convex Radius: 15mm
 - Depth: 1.5-35.0
 - Physical Footprint: 32.8mm*25mm
 - Footprint: 27.4mm*8.4mm
 - B-mode Frequencies: 2.6-6.5, 3.2-7.9, 4.7-12.8 MHz
 - Harmonic Frequencies: 6, 7, 8 MHz
 - Color Frequencies: 4.4, 5.0, 5.7 MHz
 - HR Flow: 5.0 MHz
 - PW Frequencies: 4.4, 5.0, 5.7 MHz
 - Biopsy Guide: NGB-018, multi angle, reusable
- SC5-1N
 - Application: Abdominal, Gynecology, Obstetrics, Vascular, Nerve, Musculo-skeletal, Urology, Thoracic/Pleural, Small Organ
 - Advanced Function: iScape View, Free Xros M, Color M, Smart 3D, Contrast Imaging (Abdomen), STE&STQ (Abdomen)
 - Bandwidth: 1.2-6.0 MHz
 - Number of Elements: 128
 - FOV (max): 74°
 - Convex Radius: 60 mm
 - Depth: 4.0-40.0 cm
 - Physical Footprint: 76.7 mm*28 mm
 - Footprint: 68 mm*18 mm
 - B-mode Frequencies: 1.2-3.8, 1.7-5.2, 2.0-6.0 MHz
 - Harmonic Frequencies: 4, 5, 6 MHz
 - Color Frequencies: 1.8, 2.0, 2.5, 3.0, 3.2 MHz
 - HR Flow: 3.5 MHz
 - PW Frequencies: 1.8, 2.0, 2.5, 3.0, 3.2 MHz
 - Biopsy Guide: NGB-022, multi angle, reusable; LPUBKG60 (disposable)
- C7-3
- Application: Abdominal, Gynecology, Obstetrics, Vascular
 - Advanced Function: iScape View, Free Xros M, Color M, Smart 3D, Contrast Imaging (Abdomen)
 - Bandwidth: 1.8-8.2 MHz
 - Number of Elements: 192
 - FOV (max): 70°
 - Convex Radius: 50 mm
 - Depth: 4.0-40.0 cm
 - Physical Footprint: 71 mm * 21.5 mm
 - Footprint: 60.5 mm * 12.2 mm
 - B-mode Frequencies: 1.8-5.4, 2.3-6.8, 2.8-8.2 MHz
 - Harmonic Frequencies: 4.0, 5.0, 5.5, 6.0 MHz
 - Color Frequencies: 3.0, 3.3, 3.5 MHz
 - HR Flow: 3.5 MHz
 - PW Frequencies: 3.0, 3.3, 3.6 MHz
 - Biopsy Guide: NGB-019, multi angle, reusable
- CB10-4s
 - Application: Urology
 - Advanced Function: iScape View, Free Xros M
 - Bandwidth: 2.6-12.8 MHz
 - Number of Elements: 128
 - FOV (max): 205°
 - Convex Radius: 9 mm
 - Depth: 1.5-28.0 cm
 - Physical Footprint: 22.5 mm * 20.1 mm
 - Footprint: 20.1 mm * 9 mm
 - B-mode Frequencies: 2.6-6.5, 3.2-7.9, 4.7-12.8 MHz
 - Harmonic Frequencies: 7.0, 8.0, 9.0 MHz
 - Color Frequencies: 4.4, 5.0, 5.0 MHz
 - HR Flow: 5.0 MHz
 - PW Frequencies: 4.4, 5.0, 5.7 MHz
 - Biopsy Guide: NGB-004, single angle, reusable; NGB-045, single angle, reusable
- V11-3H
 - Application: Gynecology, Obstetrics, Urology

- Advanced Function: iScape View, Free Xros M, Color M, Smart 3D, Contrast Imaging (Gynecology, Urology), Strain Elastography (Gynecology, Urology)
 - Bandwidth: 3.0-12.0 MHz
 - Number of Elements: 192
 - FOV (max): 196°
 - Convex Radius: 11 mm
 - Depth: 1.5~28.0 cm
 - Physical Footprint: 24.9 mm*21.8 mm
 - Footprint: 24 mm*9 mm
 - B-mode Frequencies: 3.0~7.0, 4.0~9.0, 5.0~12.0 MHz
 - Harmonic Frequencies: 8, 9, 10 MHz
 - Color Frequencies: 4.2, 4.4, 5.0, 5.3, 5.5 MHz
HR Flow: 5.5 MHz
 - PW Frequencies: 4.2, 4.4, 5.0, 5.3, 5.5 MHz
 - Biopsy Guide: NGB-025, single angle, reusable
- V11-3
 - Application: Gynecology, Obstetrics, Urology
 - Advanced Function: iScape View, Free Xros M, Color M, Smart 3D, Contrast Imaging (Gynecology, Urology), Strain Elastography (Gynecology, Urology)
 - Bandwidth: 3.0-12.0 MHz
 - Number of Elements: 128
 - FOV (max): 179°
 - Convex Radius: 11 mm
 - Depth: 1.5~28.0 cm
 - Physical Footprint: 24.8 mm*21.8 mm
 - Footprint: 24.1 mm*9 mm
 - B-mode Frequencies: 3.0~7.0, 4.0~9.0, 5.0~12.0 MHz
 - Harmonic Frequencies: 8, 9, 10 MHz
 - Color Frequencies: 4.4, 5.0, 5.0 MHz
HR Flow: 5.5 MHz
 - PW Frequencies: 4.4, 5.0, 5.7 MHz
 - Biopsy Guide: NGB-004, single angle, reusable; NGB-045, single angle, reusable
- V11-3B
 - Application: Gynecology, Obstetrics, Urology
 - Advanced Function: iScape View, Free Xros M, Color M, Smart 3D, Contrast Imaging (Gynecology, Urology), Strain Elastography (Gynecology, Urology)
 - Bandwidth: 3.0-12.0 MHz
 - Number of Elements: 192
 - FOV (max): 243°
 - Convex Radius: 11mm
 - Depth: 1.5-28.0 cm
 - Physical Footprint: 23.4mm*19.2mm
 - Footprint: 23.8mm*9.1mm
 - B-mode Frequencies: 3.0-7.0, 4.0-9.0, 5.0-12.0 MHz
- V11-3HB
 - Application: Gynecology, Obstetrics, Urology
 - Advanced Function: iScape View, Free Xros M, Color M, Smart 3D, Contrast Imaging (Gynecology, Urology), Strain Elastography (Gynecology, Urology)
 - Bandwidth: 3.0-12.0 MHz
 - Number of Elements: 192
 - FOV (max): 243°
 - Convex Radius: 11mm
 - Depth: 1.5-28.0 cm
 - Physical Footprint: 23.4mm*19.2mm
 - Footprint: 23.8mm*9.1mm
 - B-mode Frequencies: 3.0-7.0, 4.0-9.0, 5.0-12.0 MHz

- Harmonic Frequencies: 8, 9, 10 MHz
 - Color Frequencies: 4.4, 5.0, 5.3 MHz
 - HR Flow: 5.5 MHz
 - PW Frequencies: 4.4, 5.0, 5.3 MHz
 - Biopsy Guide: NGB-048, single angle, reusable
- 9.2 Linear array transducer**
- L13-3N
 - Application: Abdominal, Pediatric, Small Organ, Musculo-skeletal, Vascular, Nerve, Thoracic/Pleural
 - Advanced Function: iScape View, Free Xros M, Smart 3D, Contrast Imaging (Small Parts, Vascular), Strain Elastography (Small Parts, Musculo-skeletal, Vascular), STE&STQ (Small Parts, Musculo-skeletal, Vascular)
 - Bandwidth: 3.0-13.0 MHz
 - Number of Elements: 192
 - FOV (max)
 - Non-extended: 3.8cm
 - Extended: 40°
 - Steered Angle:
 - B: -12°, -6°, 0°, 6°, 12°
 - Color/PW: -30°-30°
 - Depth: 1.5-35.0 cm
 - Physical Footprint: 56.8 mm*21.2 mm
 - Footprint: 43.5 mm*8.2 mm
 - B-mode Frequencies: 3.0-9.3, 5.4-11.2, 6.6-13.0 MHz
 - Harmonic Frequencies: 9, 10, 11, 11.5, 12 MHz
 - Color Frequencies: 3.8, 4.0, 5.0, 6.2, 6.5 MHz
 - HR Flow: 7.3 MHz
 - PW Frequencies: 3.8, 4.0, 5.0, 6.2, 6.5 MHz
 - Biopsy Guide: NGB-052, multi angle, reusable; NGB-053, multi angle, reusable
 - L9-3
 - Application: Abdominal, Obstetrics, Pediatric, Small Organ, Cephalic, Musculo-skeletal, Vascular, Nerve
- Advanced Function: iScape View, Free Xros M, Smart 3D, Contrast Imaging (Small Parts, Vascular), Strain Elastography (Small Parts, Musculo-skeletal, Vascular), STE&STQ (Small Parts, Musculo-skeletal, Vascular)
 - Bandwidth: 2.5-9.0 MHz
 - Number of Elements: 192
 - FOV (max)
 - Non-extended: 4.37cm
 - Extended: 40°
 - Steered Angle:
 - B: -12°, -6°, 0°, 6°, 12°
 - Color/PW: -30°-30°
 - Depth: 1.5-35.0 cm
 - Physical Footprint: 62 mm*22 mm
 - Footprint: 48 mm*11 mm
 - B-mode Frequencies: 2.5-7.0, 3.4-8.2, 3.6-9.0 MHz
 - Harmonic Frequencies: 4.5, 5.0, 7.0 MHz
 - Color Frequencies: 3.0, 3.6, 5.0 MHz
 - HR Flow: 7.3 MHz
 - PW Frequencies: 3.0, 3.6, 5.0 MHz
 - Biopsy Guide: NGB-034, multi angle, reusable
- L16-4Hs
 - Application: Musculo-skeletal, Nerve, Small Organ, Vascular, Pediatric, Intra-operative, Cephalic
 - Advanced Function: iScape View, Free Xros M, Smart 3D, Contrast Imaging (Small Parts), Strain Elastography (Small Parts, Musculo-skeletal)
 - Bandwidth: 3.5-16 MHz
 - Number of Elements: 128
 - FOV (max)
 - Non-extended: 2.53cm
 - Extended: 40°
 - Steered Angle:
 - B: -12°, -6°, 0°, 6°, 12°
 - Color/PW: -30°-30°
 - Depth: 1.5~35.0 cm
 - Physical Footprint: 11.5 mm*38 mm
 - Footprint: 28.7 mm*5.5 mm

			<ul style="list-style-type: none"> - B-mode Frequencies: 3.5~9.3, 5.4~11.2, 9.6~16 MHz - Harmonic Frequencies: 10, 11, 12 MHz - Color Frequencies: 5.0, 6.2, 7.3 MHz - HR Flow: 6.2 MHz - PW Frequencies: 5.0, 6.2, 7.3 MHz - Biopsy Guide: Not available 		<ul style="list-style-type: none"> Imaging (Small Parts), Strain Elastography (Small Parts, Musculo-skeletal), STE&STQ (Small Parts, Musculo-skeletal) - Bandwidth: 6.0-23.0 MHz - Number of Elements: 192 - FOV (max) <ul style="list-style-type: none"> Non-extended: 2.85 cm Extended: 20° - Steered Angle: <ul style="list-style-type: none"> B: -12°, -6°, 0°, 6°, 12° Color/PW: -20°-20° - Depth: 1.5-35.0 cm - Physical Footprint: 42.23 mm * 22.10 mm - Footprint: 31.5 mm * 4.5 mm - B-mode Frequencies: 6.0-13.0, 9.0-16.6, 12.5-23.0 MHz - Harmonic Frequencies: 14.0, 16.0, 18.0 MHz - Color Frequencies: 8.9, 11.4, 13.3 MHz - HR Flow: 8.9 MHz - PW Frequencies: 8.9, 10.0, 13.3 MHz - Biopsy Guide: Not available
		9.3			Phased array transducer
				<ul style="list-style-type: none"> • SP5-1N 	<ul style="list-style-type: none"> - Application: Abdominal, Gynecology, Obstetrics, Cardiac, Pediatric, Vascular, Thoracic/ Pleural, Cephalic - Advanced Function: CW, iScape View, Free Xros M, Free Xros CM, Color M, Smart 3D, Contrast Imaging (Cardiac), TDI (Cardiac), Stress Echo (Cardiac), Tissue Tracking QA (Cardiac) - Bandwidth: 1.5-4.5 MHz - Number of Elements: 64 - FOV (max): 90° - Depth: 2.0~38.0 cm - Physical Footprint: 38.2 mm*30.4 mm - Footprint: 24 mm*15.4 mm - B-mode Frequencies: 1.5~2.5, 2.5~3.5, 3.5~4.5 MHz - Harmonic Frequencies: 3.4, 3.8, 3.8, 4.2, 4.2 MHz
			<ul style="list-style-type: none"> • L14-3W 		
			<ul style="list-style-type: none"> - Application: Abdominal, Pediatric, Small Organ, Musculo-skeletal, Vascular, Nerve, Thoracic/Pleural - Advanced Function: iScape View, Free Xros M, Smart 3D, Contrast Imaging (Small Parts, Vascular), Strain Elastography (Small Parts, Musculo-skeletal, Vascular), STE&STQ (Small Parts, Musculo-skeletal, Vascular) - Bandwidth: 3.0-14.0 MHz - Number of Elements: 256 - FOV (max) <ul style="list-style-type: none"> Non-extended: 5.8cm Extended: 20° - Steered Angle: <ul style="list-style-type: none"> B: -12°, -6°, 0°, 6°, 12° Color: -20°-20° PW: -30°-30° - Depth: 1.5~35.0 cm - Physical Footprint: 66.8 mm*25.5 mm - Footprint: 55.5 mm*8.2 mm - B-mode Frequencies: 3.0~9.0, 5.0~12.0, 6.0~14.0 MHz - Harmonic Frequencies: 10.0, 11.0, 11.5, 12.0 MHz - Color Frequencies: 5.0, 6.2, 7.2 MHz - HR Flow: 8.3 MHz - PW Frequencies: 5.0, 6.2, 7.2 MHz - Biopsy Guide: NGB-054, multi angle, reusable 		
			<ul style="list-style-type: none"> • L20-5s 		
			<ul style="list-style-type: none"> - Application: Abdominal, Pediatric, Small Organ, Musculo-skeletal, Vascular, Nerve - Advanced Function: iScape View, Free Xros M, Smart 3D, Contrast 		

- Color Frequencies: 1.9, 2.0, 2.3, 2.5, 2.6 MHz
 - HR Flow: 2.5 MHz
 - TDI: 3.0, 3.8 MHz
 - PW Frequencies: 1.9, 2.0, 2.3, 2.5, 2.6 MHz
 - TDI: 2.5, 4.0 MHz
 - CW Frequencies: 2.0 MHz
 - Biopsy Guide: NGB-011, multi angle, reusable
 - P8-2
 - Application: Abdominal, Pediatric, Cardiac, Cephalic
 - Advanced Function: CW, iScape View, Free Xros M, Free Xros CM, Color M, Smart 3D, TDI (Cardiac), Stress Echo (Cardiac), Tissue Tracking QA (Cardiac)
 - Bandwidth: 2.3-8.0 MHz
 - Number of Elements: 96
 - FOV (max): 90°
 - Depth: 2.0-38.0 cm
 - Physical Footprint: 30.5mm*23.2mm
 - Footprint: 19.5mm*11mm
 - B-mode Frequencies: 2.3-5.4, 2.8-7.4, 4.2-8.0 MHz
 - Harmonic Frequencies: 5, 6, 7 MHz
 - Color Frequencies: 2.7, 3.3, 4.0 MHz
 - HR Flow: 2.5 MHz
 - TDI: 3.0, 3.8 MHz
 - PW Frequencies: 2.7, 3.3, 4.0 MHz
 - TDI: 5.0, 6.0 MHz
 - CW: 2.5 MHz
 - Biopsy Guide: not available
 - P10-4
 - Application: Abdominal, Pediatric, Cardiac, Nerve, Cephalic
 - Advanced Function: CW, iScape View, Free Xros M, Free Xros CM, Color M, Smart 3D, TDI (Cardiac), Stress Echo (Cardiac), Tissue Tracking QA (Cardiac)
 - Bandwidth: 3.0-11.4 MHz
 - Number of Elements: 128
 - FOV (max): 90°
 - Depth: 2.0-38.0 cm
 - Physical Footprint: 15.1mm*10.2mm
 - Footprint: 15mm*9.1mm
 - B-mode Frequencies: 3.0-6.8, 3.8-10.2, 4.6-11.4 MHz
 - Harmonic Frequencies: 7.5, 8.0, 9.0 MHz
 - Color Frequencies: 4.0, 5.0, 5.3 MHz
 - HR Flow: 6.3 MHz
 - TDI: 5.7, 6.2 MHz
 - PW Frequencies: 4.0, 5.0, 5.3MHz
 - TDI: 5.0, 5.7 MHz
 - CW: 5.0 MHz
 - Biopsy Guide: not available
 - P8-2Ts
 - Application: Cardiac
 - Advanced Function: CW, Free Xros M, Free Xros CM, Color M, TDI
 - Bandwidth: 2.3-7.2MHz
 - Number of Elements: 64
 - FOV (max): 90°
 - Depth: 2.0-38.0cm
 - Physical Footprint: 14mm*12mm
 - Footprint: 10.6mm*10.6mm
 - B-mode Frequencies: 2.3-5.4, 2.8-6.4, 3.3-7.2 MHz
 - Harmonic Frequencies: 6.0, 6.5, 7.0 MHz
 - Color Frequencies: 2.7, 3.3, 4.0 MHz
 - HR Flow 2.5 MHz
 - TDI: 5.0, 6.2 MHz
 - PW Frequencies: 2.7, 3.3, 4.0 MHz
 - TDI: 2.5, 5.0 MHz
 - CW: 2.5 MHz
 - Biopsy Guide: not available
- 9.4 4D Volume transducer
- DE11-3
 - Application: Obstetrics, Gynecology, Urology
 - Advanced Function: iScape View, Free Xros M, Color M, 3D/4D, Contrast Imaging (Including Volume CEUS) (Gynecology, Urology), Strain Elastography (Gynecology, Urology)
 - Bandwidth: 2.0-9.0 MHz

	<ul style="list-style-type: none"> - Number of Elements: 128 - FOV (max): 174° - Volume Sweep Angle (max): 90° - Convex Radius: 11 mm - Depth: 4.0~40.0 cm - Physical Footprint: 24.9 mm*21.8 mm - Footprint: 24 mm*9 mm - B-mode Frequencies: 2.0~6.0, 2.8~8.2, 3.0~9.0 MHz - Harmonic Frequencies: 4, 5, 6 MHz - Color Frequencies: 4.4, 5.0, 5.7 MHz HR Flow: 5.0 MHz PW Frequencies: 4.4, 5.0, 5.7 MHz Biopsy Guide: NGB-027, single angle, reusable 		
• SD8-1		9.5	Pencil transducer
	<ul style="list-style-type: none"> - Application: Obstetrics, Gynecology, Abdominal - Advanced Function: iScape View, Free Xros M, Color M, 3D/4D - Bandwidth: 1.8-8.2 MHz - Number of Elements: 192 - FOV (max): 92° - Volume Sweep Angle (max): 85° - Convex Radius: 45 mm - Depth: 4.0~40.0 cm - Physical Footprint: 75.7 mm*52.6 mm - Footprint: 54.5 mm*14.9 mm - B-mode Frequencies: 1.8~5.4, 2.3~6.8, 2.8~8.2 MHz - Harmonic Frequencies: 4, 5, 5.5, 6 MHz - Color Frequencies: 3.0, 3.5, 4.0 MHz HR Flow: 3.5 MHz PW Frequencies: 3.0, 3.5, 3.8 MHz Biopsy Guide: NGB-039, multi angle, reusable 	9.6	Bi-Plane
• D7-2			ELC10-4
	<ul style="list-style-type: none"> - Application: Obstetrics, Gynecology, Abdominal - Advanced Function: iScape View, Free Xros M, Color M, 3D/4D - Bandwidth: 1.8-8.2 MHz - Number of Elements: 128 - FOV (max): 95° 		<ul style="list-style-type: none"> • Application: Urology • Advanced Function: iScape View, Free Xros M, Color M, Smart 3D • Bandwidth: <ul style="list-style-type: none"> - Convex: 3.0-11.0 MHz - Linear: 2.6-12.8 MHz • Number of Elements: 128 • FOV (max) <ul style="list-style-type: none"> - Convex: 164° - Linear: Non-extended: 6.58 cm; Extended: 40° • Steered Angle (Linear) <ul style="list-style-type: none"> - B: -12°, -6°, 0°, 6°, 12° - Color: -5°-5° - PW: -30°-30° • Convex Radius: 10 mm • Depth

	<ul style="list-style-type: none"> - Convex: 1.5-28.0 cm - Linear: 1.5-35.0 cm 	10.5	Gel warmer
	<ul style="list-style-type: none"> • Physical Footprint <ul style="list-style-type: none"> - Convex: 21.9 mm * 21.9 mm - Linear: 20.6 mm * 20.6 mm • Footprint <ul style="list-style-type: none"> - Convex: 21.92 mm * 11.2 mm - Linear: 72.9 mm * 11 mm • B-mode Frequencies: <ul style="list-style-type: none"> - Convex: 3.0-6.0, 4.5-7.5, 5.5-11.0 MHz - Linear: 2.6-6.5, 3.2-7.9, 4.7-12.8 MHz • Harmonic Frequencies: 7.0, 8.0, 9.0 MHz • Color Frequencies: 4.0, 5.0, 5.7 MHz 	10.6	Footswitch
	<ul style="list-style-type: none"> • HR Flow <ul style="list-style-type: none"> - Convex: 5.0 MHz - Linear: 5.7 MHz • PW Frequencies: 4.0, 5.0, 5.7 MHz • Biopsy Guide: Not available 	10.7	ECG
10	Peripheral Devices and Accessories		
10.1	Black/white video printer (digital)		
	<ul style="list-style-type: none"> • MITSUBISHI P95DW-N • SONY UP-D898MD 	10.8	PCG (not for sale in EU countries)
10.2	Black/white video printer (Digital & Analog)		
	SONY UP-X898MD		<ul style="list-style-type: none"> • PCG wave display: On/Off • Gain: 0-30, 1/step • Sweep speed: 25mm/s, 35mm/s, 50mm/s, 65mm/s, 100mm/s, 200mm/s • Invert: On/Off
10.3	Color digital video printer		
	SONY UP-D25MD		<ul style="list-style-type: none"> • Speed: 25mm/s, 35mm/s, 50mm/s, 65mm/s, 100mm/s, 200mm/s • Smooth: 1-4, 1/step
10.4	Graph/text printer	10.9	Barcode reader
	<ul style="list-style-type: none"> • Canon TS708 • Epson L130 • HP 8100(CM752A) • HP 7000(C9299A) • Epson L805 • Epson L3256 • Epson L8058 • Epson L3218 • HP DeskJet 2029 • HP DeskJet 1112 • Canon G580 • Canon G1830 • Canon G2830 		<ul style="list-style-type: none"> • SYMBOL LS2208 (1D) • SYMBOL DS4608 (2D) • JADAK HS-1M • JADAK HS-1R • Honeywell HH1800
		10.10	Built-in Wi-Fi 5 Wireless adapter
			<ul style="list-style-type: none"> • Encryption: WPA, WPA2 • Protocols: IEEE 802.11 ac/a/b/g/n • Frequency: 2.4G/5G
		10.11	iVocal Microphone
			SAMSON XPD2

10.12	Built-in Battery	745 of the European Parliament. The number adjacent to the CE marking (0123) is the code of the EU-notified body that certified meeting the requirements of Annex IX excluding CHAPTER II of the Directive.
	<ul style="list-style-type: none"> • Replaceable and rechargeable lithium battery • Full battery lasts for more than 24h in standby mode • Battery fully-recharged time: less than 4h (Under power off or standby status) • Continuous scanning time: more than 1h 	
10.13	Probe adapter	<p>PCM-SA01: transforming S socket to A socket, only for L16-4Hs, L20-5s, CB10-4s, and P8-2Ts</p> <p>NOTICE:</p> <p>Not all features or specifications described in this document may be available in all transducers and/or modes.</p>
11	System Inputs and Outputs	<p>Mindray reserves the right to make changes in specifications and features shown herein, or discontinue the product at any time without notice or obligation. Contact Mindray Representative for the most current information.</p>
11.1	Video output	<p>13 Appendix</p>
	<ul style="list-style-type: none"> • S-Video out: 1 port, PAL/NTSC • HDMI: 1 Port • VGA out: 1 port 	<p>13.1 Generic measurements</p>
11.2	Physio input	<ul style="list-style-type: none"> • Support ECG/PCG signal • ECG module: 1 port
11.3	Other input/output	<ul style="list-style-type: none"> • USB: 6 ports (2 USB 3.0 and 4 USB 2.0) • Ethernet: 1 port, 10M/100M/1000M adaption • Remote port: 1 port
12	Safety and Conformance	
12.1	Quality standards	<ul style="list-style-type: none"> • ISO 9001 • ISO 13485
12.2	Design standards	<ul style="list-style-type: none"> • EN 60601-1 and IEC 60601-1 • EN 60601-1-2 and IEC 60601-1-2 • EN 60601-1-6 and IEC 60601-1-6 • EN 60601-2-37 and IEC60601-2-37 • EN 62304 and IEC 62304 • EN 62366 and IEC 62366 • EN ISO 17664 and ISO 17664
12.3	CE declaration	<p>The ultrasound system is fully in conformance with the Regulation (EU) 2017/</p>

Color Pixel Percent(Ellipse)		HR
Color Pixel Percent(Trace)		HR(R-R)
Color Pixel Percent(Rectangle)		Time
Color Pixel Percent(Recall)		Auto Trace
Smart Trace		Manual Trace
Smart Calc		Spline Trace
-----		Acceleration
Volume		-----
Volume(Ellipse)		Ratio(Vel)
Volume(E+Dist.)		Ratio(VTI)
Ratio(D)		-----
B Ratio		Volume Flow
-----		Vas Area
Volume		TAMEAN
Volume		TAMAX
Volume(Ellipse)	13.2	Auto Calc
Volume(E+Dist.)		PS
Ratio(A)		ED
Area1		MD
Area2		PPG
Directional Ratio		TAMAX
D1		Vol Flow(TAMAX)
D2		TAMEAN
RAC		Vol Flow(TAMEAN)
Sag		Vas Diam
XS		Vas Area
Volume Flow		DT
Vas Area		MPG
TAMEAN		MMPG
TAMAX		VTI
Elas. Ratio		AT
A		S/D
B		D/S
Strain Ratio		PI
A		RI
B		PV
• M-Mode		HR
HR		13.3 Clinical option measurement package
HR(R-R)		• Abdomen
Slope		B-Mode
Distance		– Aorta Bif
Time		– Aorta Aneurysm Status
Velocity		– Shunt Diam
• D-Mode		– Portal V Diam
PS		– M Portal V Diam
ED		– Splenic V Diam
PS/ED		– PS Conflnc Diam
Vel		

- Renal V Diam
- SMV Diam
- IMV Diam
- CHD
- GB L
- GB H
- GB W
- GB wall th
- Cystic Duct
- CBD
- Panc duct
- Panc head
- Panc neck
- Panc body
- Panc tail
- Appendix
- Appendix Wall
- Pylorus
- Pylorus Wall
- Renal L
- Renal H
- Renal W
- Cortex
- Adrenal L
- Adrenal H
- Adrenal W
- Ureter
- Cortex(Renal Transplant1)
- Renal V Diam(Renal Transplant1)
- Ureter Diam(Renal Transplant1)
- Cortex(Renal Transplant2)
- Renal V Diam(Renal Transplant2)
- Ureter Diam(Renal Transplant2)
- Pre-BL L
- Pre-BL H
- Pre-BL W
- Post-BL L
- Post-BL H
- Post-BL W
- Spleen L
- Spleen H
- Spleen W
- Spleen Area
- Skin-L.Capsule Dist.
- Hepatic Lesion1 Elas.
- Hepatic Lesion2 Elas.
- Hepatic Lesion3 Elas.
- LSM
- Rt DT(Insp)
- Rt DT(Expir)
- Lt DT(Insp)
- Lt DT(Expir)
- Free Fluid
- Smart HRI
- -----
- Renal Vol
- Pre-BL Vol
- Post-BL Vol
- Mictur.Vol
- -----
- Aorta
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- Celiac Axis
- Anterior-Posterior
- Transverse
- SMA
- Anterior-Posterior
- Transverse
- C Hepatic A
- Anterior-Posterior
- Transverse
- Proper Hepatic A
- Anterior-Posterior
- Transverse
- Hepatic A
- Anterior-Posterior
- Transverse
- Pre-BL
- Transverse
- Splenic A
- Anterior-Posterior
- Transverse
- GDA
- Anterior-Posterior
- Transverse
- IMA
- Anterior-Posterior
- Transverse
- Aorta Aneurysm
- Long
- Anterior-Posterior
- Transverse
- Celiac A Aneurysm

- Long
- Anterior-Posterior
- Transverse
- SMA Aneurysm
- Long
- Anterior-Posterior
- Transverse
- C Hepatic A Aneurysm
- Long
- Anterior-Posterior
- Transverse
- Proper Hepatic A Aneurysm
- Long
- Anterior-Posterior
- Transverse
- Hepatic A Aneurysm
- Long
- Anterior-Posterior
- Transverse
- Splenic A Aneurysm
- Long
- Anterior-Posterior
- Transverse
- GDA Aneurysm
- Long
- Anterior-Posterior
- Transverse
- IMA Aneurysm
- Long
- Anterior-Posterior
- Transverse
- EVAR Residual Aneurysm Sac(2D)
- Anterior-Posterior
- Transverse
- EVAR Inflow(2D)
- Anterior-Posterior
- Transverse
- EVAR Graft Body(2D)
- Anterior-Posterior
- Transverse
- EVAR Limb(2D)
- Anterior-Posterior
- Transverse
- EVAR Outflow(2D)
- Anterior-Posterior
- Transverse
- Aortic Bypass Graft Anast(2D)
- Anterior-Posterior
- Transverse
- Aortic Bypass Graft(2D)
- Anterior-Posterior
- Transverse
- ABD Stenosis 1(2D)
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- ABD Stenosis 2(2D)
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- ABD Stenosis 3(2D)
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- ABD Stenosis 4(2D)
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- IVC
- Anterior-Posterior
- Transverse
- Checklist
- Hepatic V(2D)
- Anterior-Posterior
- Transverse
- Lt Hepatic V(2D)
- Anterior-Posterior
- Transverse
- M Hepatic V(2D)
- Anterior-Posterior
- Transverse
- Rt Hepatic V(2D)
- Anterior-Posterior
- Transverse

- Liver
- L
- H
- W
- R Liver Lobe
- L
- H
- W
- L Liver Lobe
- L
- H
- W
- Hepatic Lesion 1
- d1
- d2
- d3
- Hepatic Lesion 2
- d1
- d2
- d3
- Hepatic Lesion 3
- d1
- d2
- d3
- Hepatic Cyst 1
- d1
- d2
- d3
- Hepatic Cyst 2
- d1
- d2
- d3
- Hepatic Cyst 3
- d1
- d2
- d3
- GB
- GB L
- GB H
- GB W
- GB wall th
- GB Finding 1
- d1
- d2
- d3
- GB Finding 2
- d1
- d2
- d3
- GB Finding 3
- d1
- d2
- d3
- GB Finding 4
- d1
- d2
- d3
- GB Finding 5
- d1
- d2
- d3
- Panc Finding 1
- d1
- d2
- d3
- Panc Finding 2
- d1
- d2
- d3
- Panc Finding 3
- d1
- d2
- d3
- Panc Finding 4
- d1
- d2
- d3
- Panc Finding 5
- d1
- d2
- d3
- Kidney
- Renal L
- Renal H
- Renal W
- Cortex
- Adrenal
- Adrenal L
- Adrenal H
- Adrenal W
- Renal Lesion 1
- d1
- d2
- d3
- Renal Lesion 2
- d1

-	d2	-	L
-	d3	-	H
-	Renal Lesion 3	-	W
-	d1	-	Finding 3(Renal Transplant1)
-	d2	-	L
-	d3	-	H
-	Renal Cyst 1	-	W
-	d1	-	Finding 4(Renal Transplant1)
-	d2	-	L
-	d3	-	H
-	Renal Cyst 2	-	W
-	d1	-	Finding 5(Renal Transplant1)
-	d2	-	L
-	d3	-	H
-	Renal Cyst 3	-	W
-	d1	-	Finding 6(Renal Transplant1)
-	d2	-	L
-	d3	-	H
-	Kidney(Superior)	-	W
-	H	-	Renal Transplant 1(2D)
-	W	-	Cortex(Renal Transplant1)
-	Kidney(Mid)	-	Renal V Diam(Renal Transplant1)
-	H	-	Ureter Diam(Renal Transplant1)
-	W	-	Kidney(Renal Transplant1)
-	Kidney(Inferior)	-	L
-	H	-	H
-	W	-	W
-	Renal A	-	Adrenal(Renal Transplant1)
-	Long	-	L
-	Anterior-Posterior	-	H
-	Transverse	-	W
-	Renal A Aneurysm	-	Finding 1(Renal Transplant1)
-	Long	-	L
-	Anterior-Posterior	-	H
-	Transverse	-	W
-	Kidney(Renal Transplant1)	-	Finding 2(Renal Transplant1)
-	L	-	L
-	H	-	H
-	W	-	W
-	Adrenal(Renal Transplant1)	-	Finding 3(Renal Transplant1)
-	L	-	L
-	H	-	H
-	W	-	W
-	Finding 1(Renal Transplant1)	-	Finding 4(Renal Transplant1)
-	L	-	L
-	H	-	H
-	W	-	W
-	Finding 2(Renal Transplant1)	-	Finding 5(Renal Transplant1)

-	L	-	L
-	H	-	H
-	W	-	W
-	Finding 6(Renal Transplant1)	-	Finding 1(Renal Transplant2)
-	L	-	L
-	H	-	H
-	W	-	W
-	Kidney(Renal Transplant2)	-	Finding 2(Renal Transplant2)
-	L	-	L
-	H	-	H
-	W	-	W
-	Adrenal(Renal Transplant2)	-	Finding 3(Renal Transplant2)
-	L	-	L
-	H	-	H
-	W	-	W
-	Finding 1(Renal Transplant2)	-	Finding 4(Renal Transplant2)
-	L	-	L
-	H	-	H
-	W	-	W
-	Finding 2(Renal Transplant2)	-	Finding 5(Renal Transplant2)
-	L	-	L
-	H	-	H
-	W	-	W
-	Finding 3(Renal Transplant2)	-	Finding 6(Renal Transplant2)
-	L	-	L
-	H	-	H
-	W	-	W
-	Finding 4(Renal Transplant2)	-	Bladder
-	L	-	Pre-BL L
-	H	-	Pre-BL H
-	W	-	Pre-BL W
-	Finding 5(Renal Transplant2)	-	Post-BL L
-	L	-	Post-BL H
-	H	-	Post-BL W
-	W	-	Spleen
-	Finding 6(Renal Transplant2)	-	Spleen L
-	L	-	Spleen H
-	H	-	Spleen W
-	W	-	Spleen Area
-	Renal Transplant 2(2D)	-	Hepatic Lesion1 ElasRatio
-	Cortex(Renal Transplant2)	-	A
-	Renal V Diam(Renal Transplant2)	-	B
-	Ureter Diam(Renal Transplant2)	-	Hepatic Lesion2 ElasRatio
-	Kidney(Renal Transplant2)	-	A
-	L	-	B
-	H	-	Hepatic Lesion3 ElasRatio
-	W	-	A
-	Adrenal(Renal Transplant2)	-	B

- Smart Bladder
- XS Bladder
- Sag Bladder
- M-Mode**
- Rt DT(Insp M)
- Rt DT(Expir M)
- Lt DT(Insp M)
- Lt DT(Expir M)
- RDE(QB)
- RDE(DB)
- LDE(QB)
- LDE(DB)
- D-Mode**
- Aorta
- Celiac Axis
- SMA
- C Hepatic A
- Proper Hepatic A
- Hepatic A
- Splenic A
- GDA
- IMA
- Aorta(Post)
- Celiac Axis(Post)
- SMA(Post)
- C Hepatic A(Post)
- Proper Hepatic A(Post)
- Hepatic A(Post)
- Splenic Artery(Post)
- GDA(Post)
- IMA(Post)
- EVAR Residual Aneurysm Sac
- EVAR Inflow
- EVAR Graft Body
- EVAR Limb
- EVAR Outflow
- Aortic Bypass Graft Anast
- Aortic Bypass Graft Graft
- IVC Reflux
- IVC
- Hepatic V
- Lt Hepatic V
- M Hepatic V
- Rt Hepatic V
- Portal V
- M Portal V
- Splenic V
- Renal V
- SMV
- IMV
- Hepatic A Anast(Liver Transplant)
- Hepatic V Anast(Liver Transplant)
- Portal V Anast(Liver Transplant)
- IVC(Liver Transplant)
- Hep V Confl(Liver Transplant)
- Donor IVC(Liver Transplant)
- Renal A
- Ren A Org
- M Renal A
- Renal A1
- Renal A2
- Hilum
- Interlobar A
- Arcuate A
- Segment A
- Artery Anast(Renal Transplant1)
- Artery Anast 2(Renal Transplant1)
- Vein Anast(Renal Transplant1)
- Vein Anast 2(Renal Transplant1)
- Renal A(Renal Transplant1)
- Renal A1(Renal Transplant1)
- Renal A2(Renal Transplant1)
- Hilum(Renal Transplant1)
- Interlobar A(Renal Transplant1)
- Arcuate A(Renal Transplant1)
- Segmental A(Renal Transplant1)
- Renal Vein 1(Renal Transplant1)
- Renal Vein 2(Renal Transplant1)
- Artery Anast(Renal Transplant2)
- Artery Anast 2(Renal Transplant2)
- Vein Anast(Renal Transplant2)
- Vein Anast 2(Renal Transplant2)
- Renal A(Renal Transplant2)
- Renal A1(Renal Transplant2)
- Renal A2(Renal Transplant2)
- Hilum(Renal Transplant2)
- Interlobar A(Renal Transplant2)
- Arcuate A(Renal Transplant2)
- Segmental A(Renal Transplant2)
- Renal Vein 1(Renal Transplant2)
- Renal Vein 2(Renal Transplant2)
- TIPS
- -----
- SMA/Ao
- CA/Ao
- -----

- ABD Stenosis 1
- Pre Sten
- Sten
- Post Sten
- ABD Stenosis 2
- Pre Sten
- Sten
- Post Sten
- ABD Stenosis 3
- Pre Sten
- Sten
- Post Sten
- ABD Stenosis 4
- Pre Sten
- Sten
- Post Sten
- Renal Transplant 1(Doppler)
- Artery Anast(Renal Transplant1)
- Artery Anast 2(Renal Transplant1)
- Vein Anast(Renal Transplant1)
- Vein Anast 2(Renal Transplant1)
- Renal A(Renal Transplant1)
- Renal A1(Renal Transplant1)
- Renal A2(Renal Transplant1)
- Hilum(Renal Transplant1)
- Interlobar A(Renal Transplant1)
- Arcuate A(Renal Transplant1)
- Segmental A(Renal Transplant1)
- Renal Vein 1(Renal Transplant1)
- Renal Vein 2(Renal Transplant1)
- Renal Transplant 2(Doppler)
- Artery Anast(Renal Transplant2)
- Artery Anast 2(Renal Transplant2)
- Vein Anast(Renal Transplant2)
- Vein Anast 2(Renal Transplant2)
- Renal A(Renal Transplant2)
- Renal A1(Renal Transplant2)
- Renal A2(Renal Transplant2)
- Hilum(Renal Transplant2)
- Interlobar A(Renal Transplant2)
- Arcuate A(Renal Transplant2)
- Segmental A(Renal Transplant2)
- Renal Vein 1(Renal Transplant2)
- Renal Vein 2(Renal Transplant2)
- Gynecology
- B-Mode
- UT L
- UT H
- UT W
- Endo
- Cervix L
- Cervix H
- Cervix W
- Ovary L
- Ovary H
- Ovary W
- Follicle1 L
- Follicle1 W
- Follicle1 H
- Follicle2 L
- Follicle2 W
- Follicle2 H
- Follicle3 L
- Follicle3 W
- Follicle3 H
- Follicle4 L
- Follicle4 W
- Follicle4 H
- Follicle5 L
- Follicle5 W
- Follicle5 H
- Follicle6 L
- Follicle6 W
- Follicle6 H
- Follicle7 L
- Follicle7 W
- Follicle7 H
- Follicle8 L
- Follicle8 W
- Follicle8 H
- Follicle9 L
- Follicle9 W
- Follicle9 H
- Follicle10 L
- Follicle10 W
- Follicle10 H
- Follicle11 L
- Follicle11 W
- Follicle11 H
- Follicle12 L
- Follicle12 W
- Follicle12 H
- Follicle13 L
- Follicle13 W

- Follicle13 H
- Follicle14 L
- Follicle14 W
- Follicle14 H
- Follicle15 L
- Follicle15 W
- Follicle15 H
- Follicle16 L
- Follicle16 W
- Follicle16 H
- DWT
- BSD(R)
- BSD(Va)
- RVA(R)
- RVA(Va)
- UTA(R)
- UTA(Va)
- URA
- PVA(R)
- PVA(Va)
- PUA(R)
- PUA(Va)
- BPW-SP Dist.(R)
- BPW-SP Dist.(Va)
- Cx-SP Dist.(R)
- Cx-SP Dist.(Va)
- RA-SP Dist.(R)
- RA-SP Dist.(Va)
- Shuttle(R)
- Shuttle(Va)
- Rectocele Depth
- Intus. Depth
- ARA(R)
- ARA(Va)
- ARA(C)
- LH AP Diam(R)
- LH AP Diam(Va)
- LH AP Diam(C)
- LH Lateral Diam(R)
- LH Lateral Diam(Va)
- LH Lateral Diam(C)
- LH Area(R)
- LH Area(Va)
- LH Area(C)
- LA Angle(R)
- LA Angle(Va)
- LA Angle(C)
- LA Thickness(R)
- LA Thickness(Va)
- LA Thickness(C)
- LUG(R)
- LUG(Va)
- LUG(C)
- GYN Lesion1 Strain
- GYN Lesion2 Strain
- GYN Lesion3 Strain
- Lesion1 Elas.
- Lesion2 Elas.
- Lesion3 Elas.
- Fibroid1 Strain
- Fibroid2 Strain
- Fibroid3 Strain
- Fibroid1 Elas.
- Fibroid2 Elas.
- Fibroid3 Elas.
- -----
- UT Vol
- UT SUM
- UT-L/CX-L
- Ovary Vol
- Follicle1
- Follicle2
- Follicle3
- Follicle4
- Follicle5
- Follicle6
- Follicle7
- Follicle8
- Follicle9
- Follicle10
- Follicle11
- Follicle12
- Follicle13
- Follicle14
- Follicle15
- Follicle16
- Mean DWT
- BND
- IAS Damage
- EAS Damage
- -----
- Uterus
- UTL
- UTH
- UTW
- Endo

- Uterine Cervix
- Cervix L
- Cervix H
- Cervix W
- Fibroid 1
 - d1
 - d2
 - d3
- Fibroid 2
 - d1
 - d2
 - d3
- Fibroid 3
 - d1
 - d2
 - d3
- Uterine Finding 1
 - d1
 - d2
 - d3
- Uterine Finding 2
 - d1
 - d2
 - d3
- Uterine Finding 3
 - d1
 - d2
 - d3
- Uterine Finding 4
 - d1
 - d2
 - d3
- Uterine Finding 5
 - d1
 - d2
 - d3
- Uterine Finding 6
 - d1
 - d2
 - d3
- Ovary
 - Ovary L
 - Ovary H
 - Ovary W
- Ovarian Cyst 1
 - d1
 - d2
 - d3
- Ovarian Cyst 2
 - d1
 - d2
 - d3
- Ovarian Cyst 3
 - d1
 - d2
 - d3
- Ovarian Finding 1
 - d1
 - d2
 - d3
- Ovarian Finding 2
 - d1
 - d2
 - d3
- Ovarian Finding 3
 - d1
 - d2
 - d3
- Ovarian Finding 4
 - d1
 - d2
 - d3
- Ovarian Finding 5
 - d1
 - d2
 - d3
- Ovarian Finding 6
 - d1
 - d2
 - d3
- Follicle1
 - Follicle1 L
 - Follicle1 W
 - Follicle1 H
- Follicle2
 - Follicle2 L
 - Follicle2 W
 - Follicle2 H
- Follicle3
 - Follicle3 L
 - Follicle3 W
 - Follicle3 H
- Follicle4
 - Follicle4 L
 - Follicle4 W
 - Follicle4 H

-	Follicle5	-	GYN Lesion 1
-	Follicle5 L	-	d1
-	Follicle5 W	-	d2
-	Follicle5 H	-	d3
-	Follicle6	-	GYN Lesion 2
-	Follicle6 L	-	d1
-	Follicle6 W	-	d2
-	Follicle6 H	-	d3
-	Follicle7	-	GYN Lesion 3
-	Follicle7 L	-	d1
-	Follicle7 W	-	d2
-	Follicle7 H	-	d3
-	Follicle8	-	Residual Urine
-	Follicle8 L	-	BL Height
-	Follicle8 W	-	BL Depth
-	Follicle8 H	-	GYN Lesion1 Strain Ratio
-	Follicle9	-	A
-	Follicle9 L	-	B
-	Follicle9 W	-	GYN Lesion2 Strain Ratio
-	Follicle9 H	-	A
-	Follicle10	-	B
-	Follicle10 L	-	GYN Lesion3 Strain Ratio
-	Follicle10 W	-	A
-	Follicle10 H	-	B
-	Follicle11	-	Lesion1 Elas. Ratio
-	Follicle11 L	-	A
-	Follicle11 W	-	B
-	Follicle11 H	-	Lesion2 Elas. Ratio
-	Follicle12	-	A
-	Follicle12 L	-	B
-	Follicle12 W	-	Lesion3 Elas. Ratio
-	Follicle12 H	-	A
-	Follicle13	-	B
-	Follicle13 L	-	Fibroid1 Strain Ratio
-	Follicle13 W	-	A
-	Follicle13 H	-	B
-	Follicle14	-	Fibroid2 Strain Ratio
-	Follicle14 L	-	A
-	Follicle14 W	-	B
-	Follicle14 H	-	Fibroid3 Strain Ratio
-	Follicle15	-	A
-	Follicle15 L	-	B
-	Follicle15 W	-	Fibroid1 Elas. Ratio
-	Follicle15 H	-	A
-	Follicle16	-	B
-	Follicle16 L	-	Fibroid2 Elas. Ratio
-	Follicle16 W	-	A
-	Follicle16 H	-	B

-	Fibroid3 Elas. Ratio	-	Sac Diam1
-	A	-	Sac Diam2
-	B	-	Sac Diam3
•	Obstetrics		
	B-Mode		
-	GS	-	AF1
-	YS L	-	AF2
-	CRL	-	AF3
-	NT	-	AF4
-	BPD	-	LVIDd
-	OFD	-	LVIDs
-	HC	-	LV Diam
-	AC	-	LA Diam
-	FL	-	RVIDd
-	TAD	-	RVIDs
-	APAD	-	RV Diam
-	TCD	-	RA Diam
-	CM	-	IVSd
-	IT	-	IVSs
-	LVW	-	IVS
-	HW	-	LV Area
-	OOD	-	LA Area
-	IOD	-	RV Area
-	HUM	-	RA Area
-	Ulna	-	Ao Diam
-	RAD	-	MPA Diam
-	Tibia	-	LVOT Diam
-	FIB	-	RVOT Diam
-	CLAV	-	Facial Angle
-	Vertebrae	-	HrtA
-	MP	-	MV Diam(Z-Score)
-	Foot	-	PV Diam(Z-Score)
-	NBL	-	Ao Asc Diam(Z-Score)
-	Ear	-	Ao Desc Diam(Z-Score)
-	APTD	-	Duct Art Diam(Z-Score)
-	TTD	-	TV Diam(Z-Score)
-	FTA	-	LPA Diam(Z-Score)
-	THD	-	RPA Diam(Z-Score)
-	HrtC	-	IVC Diam(Z-Score)
-	TC	-	AV Diam(Z-Score)
-	Umb VD	-	MPA Diam(Z-Score)
-	F-kidney L	-	RV Diam(Z-Score)
-	Mat Kidney	-	LV Diam(Z-Score)
-	Cervix L	-	RV Area(Z-Score)
-	AF	-	LV Area(Z-Score)
-	NF	-	RVIDd(Z-Score)
-	Orbit	-	LVIDd(Z-Score)
-	PL Thickness	-	UT L
		-	UT H
		-	UT W

- Endo
- AH
- PH
- 3th Ventricle
- NT Above Cord
- NT Below Cord
- Mandible
- Prenasal th
- Heart AP
- Heart T
- LV Width
- LV Length
- RV Width
- RV Length
- LA Width
- RA Width
- LVWd
- LVWs
- RVWd
- RVWs
- AV Diam
- AV Area
- PV Area
- F-kidney H
- F-kidney W
- Lung
- Stomach
- YS H
- YS W
- Amniotic Sac L
- Amniotic Sac H
- Amniotic Sac W
- Ovary Cyst L
- Ovary Cyst H
- Ovary Cyst W
- UT AW
- UT PW
- CSP
- FMF
- MMF
- Lung CCAM L
- Lung CCAM H
- Lung CCAM W
- AD
- Lliac Wing Angle
- FAGL
- FAG
- Intestinum Crassum
- Liver Length
- Rib Length
- Shoulder Blade
- -----
- MAD
- Mean Sac Diam
- AFI
- EFW
- EFW2
- HC/AC(Campbell)
- FL/AC
- FL/BPD
- AXT
- CI
- FL/HC(Hadlock)
- AC(c)
- HC(c)
- HrtC/TC
- TCD/AC
- LVW/HW
- LVD/RVD
- LAD/RAD
- AoD/MPAD
- LAD/AoD
- UT Vol
- UT SUM
- UT-L/CX-L
- -----
- AFI
- AF1
- AF2
- AF3
- AF4
- Uterus
- UTL
- UTH
- UTW
- Endo
- M-Mode**
- FHR (M)
- LVIDd
- LVIDs
- RVIDd
- RVIDs
- IVSd
- IVSs
- RVIDd(Z-Score)
- LVIDd(Z-Score)

- MVE
- TVE
- AVE
- MAPSE
- TAPSE
- LV ICT
- LV IRT
- LV ET
- RV ICT
- RV IRT
- RV ET
- D-Mode**
- Umb A
- Duct Veno
- Placenta A
- MCA
- Fetal Ao
- Desc Aorta
- Ut A
- Ovarian A
- FHR (Doppler)
- Asc Aorta
- RVOT
- LVOT
- MV E
- MV A
- TV E
- TV A
- MV E'
- MV A'
- MV S'
- TV E'
- TV A'
- TV S'
- AV PV
- AV VTI
- PV PV
- PV VTI
- Duct Art PV
- Duct Art VTI
- AV TPV
- PV TPV
- Duct Art TPV
- Thoracic Aorta
- Hepatic Vein
- IVC
- Umb V
- Ovary
- Endometrium
- Cervical Cancer
- Fibroid
- Duct Art
- ICA
- Celiac A
- -----
- MV E/A
- TV E/A
- MV E/E'
- TV E/E'
- **Cardiology**
- B-Mode**
- RVAWd(2D)
- RVAWs(2D)
- RVDd(2D)
- RVDs(2D)
- IVSd(2D)
- IVSs(2D)
- LVIDd(2D)
- LVIDs(2D)
- LVPWd(2D)
- LVPWs(2D)
- Diastole(2D)
- Systole(2D)
- LVd apical
- LVLs apical
- LVAd apical
- LVAs apical
- LVAd sax MV
- LVAs sax MV
- LVAd sax Endo
- LVAd sax Epi
- LV Major
- LV Minor
- LV Area(d)
- LV Area(s)
- HR(2D)
- RA Major
- RA Minor
- RA Area
- RA Vol(A4C)
- RAP
- RV Area(d)
- RV Area(s)
- RV Major
- RV Minor
- LA Diam(2D)

- LA Major
- LA Minor
- LA Area
- LVOT Diam
- Ao Diam(2D)
- ACS(2D)
- AV Diam
- Ao Isthmus(2D)
- Ao Sinus Diam(2D)
- Ao st junct(2D)
- AVA
- Ao Arch Diam(2D)
- Ao Asc Diam(2D)
- Ao Desc Diam(2D)
- Duct Art Diam
- Post Ductal
- Pre Ductal
- MCS(2D)
- MV Diam
- MV EPSS(2D)
- MVA
- TV Diam
- TVA
- PV Diam
- RVOT Diam
- MPA Diam(2D)
- RPA Diam(2D)
- LPA Diam(2D)
- IVC Diam(Expir)
- IVC Diam(Insp)
- SVC Diam(Expir)
- SVC Diam(Insp)
- LCA Diam
- RCA Diam
- PEd(2D)
- PEs(2D)
- VSD Diam
- ASD Diam
- PDA Diam
- PFO Diam
- AutoEF
- Rt DT(Insp)
- Rt DT(Expir)
- Lt DT(Insp)
- Lt DT(Expir)
- -----
- LA/Ao(2D)
- -----
- LV(2D)
- Diastole(2D)
- Systole(2D)
- IVSd(2D)
- LVIDd(2D)
- LVPWd(2D)
- IVSs(2D)
- LVIDs(2D)
- LVPWs(2D)
- HR(2D)
- Simpson
- A4Cd
- A4Cs
- A2Cd
- A2Cs
- HR(2D)
- Mod.Simpson
- LVLD apical
- LVLs apical
- LVAd sax MV
- LVAs sax MV
- LVAd sax PM
- LVAs sax PM
- HR(2D)
- S-P Ellipse
- LVLD apical
- LVAd apical
- LVLs apical
- LVAs apical
- HR(2D)
- B-P Ellipse
- LVIDd(2D)
- LVAd sax MV
- LVIDs(2D)
- LVAs sax MV
- LVAd apical
- LVAs apical
- HR(2D)
- Bullet
- LVLD apical
- LVLs apical
- LVAd sax MV
- LVAs sax MV
- HR(2D)
- LV Mass(Cube-2D)
- IVSd(2D)
- LVIDd(2D)
- LVPWd(2D)

- LV Mass(A-L)
- LVd apical
- LVAd sax Epi
- LVAd sax Endo
- LV Mass(T-E)
- LVAd sax Epi
- LVAd sax Endo
- a
- d
- LA Vol(Simp)
- LA Vol(A2C)
- LA Vol(A4C)
- LA Vol(A-L)
- LA apical
- LAA(A2C)
- LAA(A4C)
- MVA(VTI)
- LVOT Diam
- LVOT VTI
- MV VTI
- AVA(VTI)
- LVOT Diam
- LVOT VTI
- AV VTI
- CO(LVOT)
- LVOT Diam
- LVOT VTI
- AV HR
- CO(RVOT)
- RVOT Diam
- RVOT VTI
- PV HR
- CO(MV)
- MV Diam
- MV VTI
- MV HR
- CO(TV)
- TV Diam
- TV VTI
- TV HR
- PISA MR
- MR Rad
- MR Als Vel
- MR VTI
- PISA AR
- AR Rad
- AR Als Vel
- AR VTI
- PISA TR
- TR Rad
- TR Als Vel
- TR VTI
- PISA PR
- PR Rad
- PR Als Vel
- PR VTI
- Qp/Qs
- LVOT Diam
- LVOT VTI
- RVOT Diam
- RVOT VTI
- Z-Scores ($\leq 3Y$) (2D)
- AV Diam
- Ao Sinus Diam
- Ao st junct
- PV Diam
- Ao Arch IA-LCA
- Ao Arch LCA-LSA
- Ao Arch after LSA
- Ao Isthmus
- Thoracic Ao Diam
- IVC Diam
- MV Diam
- TV Diam
- MPA Diam
- RPA Diam
- LPA Diam
- Z-Scores ($< 18Y$) (2D)
- LV Area(d) A4C
- LV Area(s) A4C
- LVIDd A4C(2D)
- LVIDs A4C(2D)
- LA AP Diam A4C
- LA LL Diam A4C
- LA Area A4C
- RA AP Diam A4C
- RA LL Diam A4C
- RA Area A4C
- RV Area(d) A4C
- RV Area(s) A4C
- RVd Major A4C
- RVs Major A4C
- RVd Minor (basal) A4C
- RVd Minor (midcavity) A4C
- LV Area(d) A2C
- LV Area(s) A2C

- LVIDd A2C(2D)
- LVIDs A2C(2D)
- M-Mode**
- RVAWd(M)
- RVAWS(M)
- RVDd(M)
- RVDs(M)
- Ao Arch Diam(M)
- Ao Asc Diam(M)
- Ao Desc Diam(M)
- Ao Diam(M)
- Ao Isthmus(M)
- Ao Sinus Diam(M)
- Ao st junct(M)
- ACS(M)
- HR(M)
- IVSd(M)
- IVSs(M)
- LA Diam(M)
- LPA Diam(M)
- Diastole(M)
- Systole(M)
- LVET(M)
- LVIDd(M)
- LVIDs(M)
- LVOT Diam
- LVPEP(M)
- LVPWd(M)
- LVPWs(M)
- MCS(M)
- MPA Diam(M)
- MV A Amp
- MV E Amp
- MV D-E Slope
- MV D-E Amp
- MV E-F Slope
- MV EPSS(M)
- PEd(M)
- PEs(M)
- RPA Diam(M)
- RVET(M)
- RVOT Diam
- RVPEP(M)
- MAPSE
- TAPSE
- MV ALL
- IVC Diam(Insp)(M)
- IVC Diam(Expir)(M)
- SVC Diam(Insp)(M)
- SVC Diam(Expir)(M)
- Rt DT(Insp M)
- Rt DT(Expir M)
- Lt DT(Insp M)
- Lt DT(Expir M)
- RDE(QB)
- RDE(DB)
- LDE(QB)
- LDE(DB)
- -----
- LA/Ao(M)
- -----
- LV(M)
- Diastole(M)
- Systole(M)
- IVSd(M)
- LVIDd(M)
- LVPWd(M)
- IVSs(M)
- LVIDs(M)
- LVPWs(M)
- HR(M)
- LV Mass(Cube-M)
- IVSd(M)
- LVIDd(M)
- LVPWd(M)
- LV Tei Index(M)
- MV C-O dur(M)
- LVET(M)
- Z-Scores (??3Y) (M)
- IVSd(M)
- LVPWd(M)
- Z-Scores (<18Y) (M)
- LVIDd(M)
- LVIDs(M)
- D-Mode**
- MV Aa(lateral)
- MV Aa(medial)
- AAo Vmax
- AV VTI
- AV HR
- AV Vmax
- AR DecT
- AR Time
- AR PHT
- AR Ved
- AR Vmax

- AR VTI
- MV ARa(lateral)
- MV ARa(medial)
- ASD Vmax
- AV AccT
- AV DecT
- Coarc Post-Duct
- Coarc Pre-Duct
- DAo Vmax
- MV DRa(lateral)
- MV DRa(medial)
- MV Ea(lateral)
- MV Ea(medial)
- IVC Vel(Expir)
- IVC Vel(Insp)
- IVCT
- LPA Vmax
- LVET(Doppler)
- LVOT AccT
- LVOT VTI
- LVOT Vmax
- LVPEP(Doppler)
- MPA Vmax
- dP/dt
- Tau(BAI)
- MR VTI
- MR Vmax
- MS Vmax
- MV A Dur
- MV A Vel
- MV A VTI
- MV AccT
- MV DecT
- MV E Dur
- MV E Vel
- MV E VTI
- IVRT
- MV VTI
- MV HR
- MV Vmax
- PVein A Dur
- PVein A Vel
- PVein D Vel
- PVein D VTI
- PVein DecT
- PVein S Vel
- PVein S VTI
- PDA Vel(d)
- PDA Vel(s)
- PR PHT
- PR VTI
- PR Ved
- PR Vmax
- PR DecT
- PV AccT
- PV VTI
- PV HR
- PV Vmax
- RAP
- RPA Vmax
- RVET(Doppler)
- RVOT Vmax
- RVOT VTI
- RVPEP(Doppler)
- MV Sa(lateral)
- MV Sa(medial)
- SVC Vel(Expir)
- SVC Vel(Insp)
- TR VTI
- TR Vmax
- TV A Dur
- TV A Vel
- TV AccT
- TV DecT
- TV E Vel
- TV VTI
- TV HR
- TV Vmax
- VSD Vmax
- Hepatic V S Vel
- Hepatic V D Vel
- -----
- MV E/A
- MVA(PHT)
- TV E/A
- -----
- LV Tei Index(Doppler)
- MV C-O dur(Doppler)
- LVET(Doppler)
- RVSP
- TR Vmax
- RAP
- PAEDP
- PR Ved
- RAP

- MVA(VTI)
- LVOT Diam
- LVOT VTI
- MV VTI
- AVA(VTI)
- LVOT Diam
- LVOT VTI
- AV VTI
- CO(LVOT)
- LVOT Diam
- LVOT VTI
- AV HR
- CO(RVOT)
- RVOT Diam
- RVOT VTI
- PV HR
- CO(MV)
- MV Diam
- MV VTI
- MV HR
- CO(TV)
- TV Diam
- TV VTI
- TV HR
- RV Tei Index
- TV C-O dur
- RVET(Doppler)
- PISA MR
- MR Rad
- MR Als Vel
- MR VTI
- PISA AR
- AR Rad
- AR Als Vel
- AR VTI
- PISA TR
- TR Rad
- TR Als Vel
- TR VTI
- PISA PR
- PR Rad
- PR Als Vel
- PR VTI
- Qp/Qs
- LVOT Diam
- LVOT VTI
- RVOT Diam
- RVOT VTI
- Urology
- B-Mode
- Renal L
- Renal H
- Renal W
- Cortex
- Adrenal L
- Adrenal H
- Adrenal W
- Ureter
- Cortex(Renal Transplant1)
- Renal V Diam(Renal Transplant1)
- Ureter Diam(Renal Transplant1)
- Cortex(Renal Transplant2)
- Renal V Diam(Renal Transplant2)
- Ureter Diam(Renal Transplant2)
- Prostate L
- Prostate H
- Prostate W
- Seminal L
- Seminal H
- Seminal W
- Urethra
- Pre-BL L
- Pre-BL H
- Pre-BL W
- Post-BL L
- Post-BL H
- Post-BL W
- Testicular L
- Testicular H
- Testicular W
- Epididymis L
- Epididymis H
- Epididymis W
- Scrotal Wall
- Testis V(2D)
- Testis V(Valsalva 2D)
- Prostate Mass1 Strain
- Prostate Mass2 Strain
- Prostate Mass3 Strain
- Prostate Mass1 Elas.
- Prostate Mass2 Elas.
- Prostate Mass3 Elas.
- -----
- Renal Vol
- Prostate Vol
- Pre-BL Vol

- Post-BL Vol
- Mictur.Vol
- Testicular Vol
- -----
- Kidney
 - Renal L
 - Renal H
 - Renal W
 - Cortex
 - Adrenal
 - Adrenal L
 - Adrenal H
 - Adrenal W
- Renal Lesion 1
 - d1
 - d2
 - d3
- Renal Lesion 2
 - d1
 - d2
 - d3
- Renal Lesion 3
 - d1
 - d2
 - d3
- Renal Cyst 1
 - d1
 - d2
 - d3
- Renal Cyst 2
 - d1
 - d2
 - d3
- Renal Cyst 3
 - d1
 - d2
 - d3
- Kidney(Superior)
 - H
 - W
- Kidney(Mid)
 - H
 - W
- Kidney(Inferior)
 - H
 - W
- Renal A
- Long
- Anterior-Posterior
- Transverse
- Renal A Aneurysm
- Long
- Anterior-Posterior
- Transverse
- Kidney(Renal Transplant1)
- L
- H
- W
- Adrenal(Renal Transplant1)
- L
- H
- W
- Finding 1(Renal Transplant1)
- L
- H
- W
- Finding 2(Renal Transplant1)
- L
- H
- W
- Finding 3(Renal Transplant1)
- L
- H
- W
- Finding 4(Renal Transplant1)
- L
- H
- W
- Finding 5(Renal Transplant1)
- L
- H
- W
- Finding 6(Renal Transplant1)
- L
- H
- W
- Renal Transplant 1(2D)
- Cortex(Renal Transplant1)
- Renal V Diam(Renal Transplant1)
- Ureter Diam(Renal Transplant1)
- Kidney(Renal Transplant1)
- L
- H
- W
- Adrenal(Renal Transplant1)
- L

- H
- W
- Finding 1(Renal Transplant1)
 - L
 - H
 - W
- Finding 2(Renal Transplant1)
 - L
 - H
 - W
- Finding 3(Renal Transplant1)
 - L
 - H
 - W
- Finding 4(Renal Transplant1)
 - L
 - H
 - W
- Finding 5(Renal Transplant1)
 - L
 - H
 - W
- Finding 6(Renal Transplant1)
 - L
 - H
 - W
- Kidney(Renal Transplant2)
 - L
 - H
 - W
- Adrenal(Renal Transplant2)
 - L
 - H
 - W
- Finding 1(Renal Transplant2)
 - L
 - H
 - W
- Finding 2(Renal Transplant2)
 - L
 - H
 - W
- Finding 3(Renal Transplant2)
 - L
 - H
 - W
- Finding 4(Renal Transplant2)
 - L
 - H
 - W
- Finding 5(Renal Transplant2)
 - L
 - H
 - W
- Finding 6(Renal Transplant2)
 - L
 - H
 - W
- Prostate
- Prostate L

- Prostate H
 - Prostate W
 - Prostate2
 - Long
 - Anterior-Posterior
 - Coronal
 - Seminal Vesicle
 - Seminal L
 - Seminal H
 - Seminal W
 - Prostate Mass 1
 - d1
 - d2
 - d3
 - Prostate Mass 2
 - d1
 - d2
 - d3
 - Prostate Mass 3
 - d1
 - d2
 - d3
 - Bladder
 - Pre-BL L
 - Pre-BL H
 - Pre-BL W
 - Post-BL L
 - Post-BL H
 - Post-BL W
 - Smart Bladder
 - XS Bladder
 - Sag Bladder
 - Testis
 - Testicular L
 - Testicular H
 - Testicular W
 - Testis Mass 1
 - d1
 - d2
 - d3
 - Testis Mass 2
 - d1
 - d2
 - d3
 - Testis Mass 3
 - d1
 - d2
 - d3
- Epididymis
 - Epididymis L
 - Epididymis H
 - Epididymis W
 - Testicle(Superior)
 - H
 - W
 - Testicle(Mid)
 - H
 - W
 - Testicle(Inferior)
 - H
 - W
 - Epididymal Head
 - L
 - H
 - W
 - Epididymal Body
 - L
 - H
 - W
 - Epididymal Tail
 - L
 - H
 - W
 - Prostate Mass1 Strain Ratio
 - A
 - B
 - Prostate Mass2 Strain Ratio
 - A
 - B
 - Prostate Mass3 Strain Ratio
 - A
 - B
 - Prostate Mass1 Elas. Ratio
 - A
 - B
 - Prostate Mass2 Elas. Ratio
 - A
 - B
 - Prostate Mass3 Elas. Ratio
 - A
 - B
 - D-Mode
 - Renal A
 - Ren A Org
 - M Renal A
 - Renal A1

- Renal A2
 - Hilum
 - Interlobar A
 - Arcuate A
 - Segment A
 - Artery Anast(Renal Transplant1)
 - Artery Anast 2(Renal Transplant1)
 - Vein Anast(Renal Transplant1)
 - Vein Anast 2(Renal Transplant1)
 - Renal A(Renal Transplant1)
 - Renal A1(Renal Transplant1)
 - Renal A2(Renal Transplant1)
 - Hilum(Renal Transplant1)
 - Interlobar A(Renal Transplant1)
 - Arcuate A(Renal Transplant1)
 - Segmental A(Renal Transplant1)
 - Renal Vein 1(Renal Transplant1)
 - Renal Vein 2(Renal Transplant1)
 - Artery Anast(Renal Transplant2)
 - Artery Anast 2(Renal Transplant2)
 - Vein Anast(Renal Transplant2)
 - Vein Anast 2(Renal Transplant2)
 - Renal A(Renal Transplant2)
 - Renal A1(Renal Transplant2)
 - Renal A2(Renal Transplant2)
 - Hilum(Renal Transplant2)
 - Interlobar A(Renal Transplant2)
 - Arcuate A(Renal Transplant2)
 - Segmental A(Renal Transplant2)
 - Renal Vein 1(Renal Transplant2)
 - Renal Vein 2(Renal Transplant2)
 - Testis A
 - Testis V
 - Testis V(Valsalva)
 - Epididymis A
 - Epididymis V
 - -----
 - Renal Transplant 1(Doppler)
 - Artery Anast(Renal Transplant1)
 - Artery Anast 2(Renal Transplant1)
 - Vein Anast(Renal Transplant1)
 - Vein Anast 2(Renal Transplant1)
 - Renal A(Renal Transplant1)
 - Renal A1(Renal Transplant1)
 - Renal A2(Renal Transplant1)
 - Hilum(Renal Transplant1)
 - Interlobar A(Renal Transplant1)
 - Arcuate A(Renal Transplant1)
 - Segmental A(Renal Transplant1)
 - Renal Vein 1(Renal Transplant1)
 - Renal Vein 2(Renal Transplant1)
 - Artery Anast(Renal Transplant2)
 - Artery Anast 2(Renal Transplant2)
 - Vein Anast(Renal Transplant2)
 - Vein Anast 2(Renal Transplant2)
 - Renal A(Renal Transplant2)
 - Renal A1(Renal Transplant2)
 - Renal A2(Renal Transplant2)
 - Hilum(Renal Transplant2)
 - Interlobar A(Renal Transplant2)
 - Arcuate A(Renal Transplant2)
 - Segmental A(Renal Transplant2)
 - Renal Vein 1(Renal Transplant2)
 - Renal Vein 2(Renal Transplant2)
- Vascular
- B-Mode
- CCA IMT
 - Bulb IMT
 - ICA IMT
 - ECA IMT
 - -----
 - IMT
 - CCA IMT
 - Bulb IMT
 - ICA IMT
 - ECA IMT
 - CCA
 - Anterior-Posterior
 - Transverse
 - Outer Diameter
 - Inner Diameter
 - Outer Area
 - Inner Area
 - Bulb
 - Anterior-Posterior
 - Transverse
 - Outer Diameter
 - Inner Diameter
 - Outer Area
 - Inner Area
 - Carotid Bifurcation
 - Anterior-Posterior
 - Transverse

- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- ICA
 - Anterior-Posterior
 - Transverse
 - Outer Diameter
 - Inner Diameter
 - Outer Area
 - Inner Area
- ECA
 - Anterior-Posterior
 - Transverse
 - Outer Diameter
 - Inner Diameter
 - Outer Area
 - Inner Area
- Vert A
 - Anterior-Posterior
 - Transverse
 - Outer Diameter
 - Inner Diameter
 - Outer Area
 - Inner Area
- Subclav A
 - Anterior-Posterior
 - Transverse
 - Outer Diameter
 - Inner Diameter
 - Outer Area
 - Inner Area
- Innom A
 - Anterior-Posterior
 - Transverse
 - Outer Diameter
 - Inner Diameter
 - Outer Area
 - Inner Area
- Mammary A
 - Anterior-Posterior
 - Transverse
 - Outer Diameter
 - Inner Diameter
 - Outer Area
 - Inner Area
- CCA Aneurysm
- Long
- Anterior-Posterior
- Transverse
- Bulb Aneurysm
- Long
- Anterior-Posterior
- Transverse
- Carotid Bifurcation Aneurysm
- Long
- Anterior-Posterior
- Transverse
- ICA Aneurysm
- Long
- Anterior-Posterior
- Transverse
- ECA Aneurysm
- Long
- Anterior-Posterior
- Transverse
- Vert A Aneurysm
- Long
- Anterior-Posterior
- Transverse
- Subclav A Aneurysm
- Long
- Anterior-Posterior
- Transverse
- Innom A Aneurysm
- Long
- Anterior-Posterior
- Transverse
- Mammary A Aneurysm
- Long
- Anterior-Posterior
- Transverse
- Carotid Graft 1 Anast
- Long
- Anterior-Posterior
- Transverse
- Carotid Graft 1 Graft
- Long
- Anterior-Posterior
- Transverse
- Carotid Graft 2 Anast
- Long
- Anterior-Posterior
- Transverse
- Carotid Graft 2 Graft
- Long

- Anterior-Posterior
- Transverse
- Carotid Graft 3 Anast
- Long
- Anterior-Posterior
- Transverse
- Carotid Graft 3 Graft
- Long
- Anterior-Posterior
- Transverse
- Carotid Stent 1
- Long
- Anterior-Posterior
- Transverse
- Carotid Stent 2
- Long
- Anterior-Posterior
- Transverse
- Carotid Stent 3
- Long
- Anterior-Posterior
- Transverse
- Carotid Stenosis 1
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- Carotid Stenosis 2
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- Carotid Stenosis 3
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- Carotid Stenosis 4
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- Axill A
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- Brachial A
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- Radial A
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- Ulnar A
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- Axill A Aneurysm
- Long
- Anterior-Posterior
- Transverse
- Brachial A Aneurysm
- Long
- Anterior-Posterior
- Transverse
- Radial A Aneurysm
- Long
- Anterior-Posterior
- Transverse
- Ulnar A Aneurysm
- Long
- Anterior-Posterior
- Transverse
- UE A Graft 1 Native Inflow
- Anterior-Posterior

- Transverse
- UE A Graft 1 Anast
- Anterior-Posterior
- Transverse
- UE A Graft 1 Graft
- Anterior-Posterior
- Transverse
- UE A Graft 1 Native Outflow
- Anterior-Posterior
- Transverse
- UE A Graft 2 Native Inflow
- Anterior-Posterior
- Transverse
- UE A Graft 2 Anast
- Anterior-Posterior
- Transverse
- UE A Graft 2 Graft
- Anterior-Posterior
- Transverse
- UE A Graft 2 Native Outflow
- Anterior-Posterior
- Transverse
- UE A Graft 3 Native Inflow
- Anterior-Posterior
- Transverse
- UE A Graft 3 Anast
- Anterior-Posterior
- Transverse
- UE A Graft 3 Graft
- Anterior-Posterior
- Transverse
- UE A Graft 3 Native Outflow
- Anterior-Posterior
- Transverse
- UE A Stent 1
- Long
- Anterior-Posterior
- Transverse
- UE A Stent 2
- Long
- Anterior-Posterior
- Transverse
- UE A Stent 3
- Long
- Anterior-Posterior
- Transverse
- UE A Stenosis 1(2D)
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- UE A Stenosis 2(2D)
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- UE A Stenosis 3(2D)
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- UE A Stenosis 4(2D)
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- C.Iliac A
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- Ex.Iliac A
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- IIA
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- CFA

- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- DFA
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- SFA
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- Pop A
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- TP Trunk A
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- A.Tib A
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- Peroneal A
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- P.Tib A
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- Dors.Ped. A
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- C.Iliac A Aneurysm
- Long
- Anterior-Posterior
- Transverse
- Ex.Iliac A Aneurysm
- Long
- Anterior-Posterior
- Transverse
- IIA Aneurysm
- Long
- Anterior-Posterior
- Transverse
- CFA Aneurysm
- Long
- Anterior-Posterior
- Transverse
- DFA Aneurysm
- Long
- Anterior-Posterior
- Transverse
- SFA Aneurysm
- Long
- Anterior-Posterior
- Transverse
- Pop A Aneurysm
- Long
- Anterior-Posterior
- Transverse
- TP Trunk A Aneurysm
- Long
- Anterior-Posterior
- Transverse
- A.Tib A Aneurysm
- Long

- Anterior-Posterior
- Transverse
- Peroneal A Aneurysm
- Long
- Anterior-Posterior
- Transverse
- P.Tib A Aneurysm
- Long
- Anterior-Posterior
- Transverse
- Dors.Ped. A Aneurysm
- Long
- Anterior-Posterior
- Transverse
- LE A Graft 1 Native Inflow
- Anterior-Posterior
- Transverse
- LE A Graft 1 Anast
- Anterior-Posterior
- Transverse
- LE A Graft 1 Graft
- Anterior-Posterior
- Transverse
- LE A Graft 1 Native Outflow
- Anterior-Posterior
- Transverse
- LE A Graft 2 Native Inflow
- Anterior-Posterior
- Transverse
- LE A Graft 2 Anast
- Anterior-Posterior
- Transverse
- LE A Graft 2 Graft
- Anterior-Posterior
- Transverse
- LE A Graft 2 Native Outflow
- Anterior-Posterior
- Transverse
- LE A Graft 3 Native Inflow
- Anterior-Posterior
- Transverse
- LE A Graft 3 Anast
- Anterior-Posterior
- Transverse
- LE A Graft 3 Graft
- Anterior-Posterior
- Transverse
- LE A Graft 3 Native Outflow
- Anterior-Posterior
- Transverse
- LE A Stent 1
- Long
- Anterior-Posterior
- Transverse
- LE A Stent 2
- Long
- Anterior-Posterior
- Transverse
- LE A Stent 3
- Long
- Anterior-Posterior
- Transverse
- LE A Stenosis 1(2D)
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- LE A Stenosis 2(2D)
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- LE A Stenosis 3(2D)
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- LE A Stenosis 4(2D)
- Anterior-Posterior
- Transverse
- Outer Diameter
- Inner Diameter
- Outer Area
- Inner Area
- LE A Finding 1
- Long
- Anterior-Posterior
- Transverse
- LE A Finding 2
- Long

- Anterior-Posterior
- Transverse
- LE A Finding 3
- Long
- Anterior-Posterior
- Transverse
- LE A Finding 4
- Long
- Anterior-Posterior
- Transverse
- LE A Finding 5
- Long
- Anterior-Posterior
- Transverse
- LE A Finding 6
- Long
- Anterior-Posterior
- Transverse
- Int Jug V
- Anterior-Posterior
- Transverse
- Checklist
- Innom V
- Anterior-Posterior
- Transverse
- Checklist
- Subclav V
- Anterior-Posterior
- Transverse
- Checklist
- Ax V
- Anterior-Posterior
- Transverse
- Checklist
- Brachial V
- Anterior-Posterior
- Transverse
- Checklist
- Radial V
- Anterior-Posterior
- Transverse
- Checklist
- Ulnar V
- Anterior-Posterior
- Transverse
- Checklist
- Volar V
- Anterior-Posterior
- Transverse
- Checklist
- Cephalic V
- Anterior-Posterior
- Transverse
- Checklist
- Basilic V
- Anterior-Posterior
- Transverse
- Checklist
- CA Junction
- Anterior-Posterior
- Transverse
- Checklist
- Upper Arm Cephalic V
- Anterior-Posterior
- Transverse
- Checklist
- Cephalic-Antecubital V
- Anterior-Posterior
- Transverse
- Checklist
- Forearm Cephalic V
- Anterior-Posterior
- Transverse
- Checklist
- BA Junction
- Anterior-Posterior
- Transverse
- Checklist
- Upper Arm Basilic V
- Anterior-Posterior
- Transverse
- Checklist
- Basilic-Antecubital V
- Anterior-Posterior
- Transverse
- Checklist
- Forearm Basilic V
- Anterior-Posterior
- Transverse
- Checklist
- Digital V
- Anterior-Posterior
- Transverse
- Checklist
- Median Cubital V
- Anterior-Posterior

- Transverse
- Checklist
- AVF-Inflow Artery
- Anterior-Posterior
- Transverse
- AVF-Anast
- Anterior-Posterior
- Transverse
- AVF-Outflow Vein Level 1
- Anterior-Posterior
- Transverse
- AVF-Outflow Vein Level 2
- Anterior-Posterior
- Transverse
- AVF-Outflow Vein Level 3
- Anterior-Posterior
- Transverse
- AVF-Outflow Vein Level 4
- Anterior-Posterior
- Transverse
- AVF-Outflow Vein Level 5
- Anterior-Posterior
- Transverse
- AVF-Outflow Vein Level 6
- Anterior-Posterior
- Transverse
- AVF-Stenosis 1
- Anterior-Posterior
- Transverse
- AVF-Stenosis 2
- Anterior-Posterior
- Transverse
- AVF-Stenosis 3
- Anterior-Posterior
- Transverse
- AVF-Aneurysm 1
- Anterior-Posterior
- Transverse
- AVF-Aneurysm 2
- Anterior-Posterior
- Transverse
- AVF-Aneurysm 3
- Anterior-Posterior
- Transverse
- AV Graft-Inflow Artery
- Anterior-Posterior
- Transverse
- AV Graft-Arterial Anast
- Anterior-Posterior
- Transverse
- AV Graft-Graft
- Anterior-Posterior
- Transverse
- AV Graft-Venous Anast
- Anterior-Posterior
- Transverse
- AV Graft-Outflow Vein Level 1
- Anterior-Posterior
- Transverse
- AV Graft-Outflow Vein Level 2
- Anterior-Posterior
- Transverse
- AV Graft-Outflow Vein Level 3
- Anterior-Posterior
- Transverse
- AV Graft-Outflow Vein Level 4
- Anterior-Posterior
- Transverse
- AV Graft-Outflow Vein Level 5
- Anterior-Posterior
- Transverse
- AV Graft-Outflow Vein Level 6
- Anterior-Posterior
- Transverse
- C.Iliac V
- Anterior-Posterior
- Transverse
- Checklist
- Ex.Iliac V
- Anterior-Posterior
- Transverse
- Checklist
- IIV
- Anterior-Posterior
- Transverse
- Checklist
- CFV
- Anterior-Posterior
- Transverse
- Checklist
- Femoral V
- Anterior-Posterior
- Transverse
- Checklist
- DFV
- Anterior-Posterior

- Transverse
- Checklist
- Pop V
- Anterior-Posterior
- Transverse
- Checklist
- P.Tib V
- Anterior-Posterior
- Transverse
- Checklist
- Peroneal V
- Anterior-Posterior
- Transverse
- Checklist
- Sural V
- Anterior-Posterior
- Transverse
- Checklist
- Soleal V
- Anterior-Posterior
- Transverse
- Checklist
- A.Tib V
- Anterior-Posterior
- Transverse
- Checklist
- TP Trunk V
- Anterior-Posterior
- Transverse
- Checklist
- Saph V
- Anterior-Posterior
- Transverse
- Checklist
- SSV
- Anterior-Posterior
- Transverse
- Checklist
- SF Junction
- Anterior-Posterior
- Transverse
- Checklist
- GSV Thigh
- Anterior-Posterior
- Transverse
- Checklist
- GSV Knee
- Anterior-Posterior
- Transverse
- Checklist
- GSV Calf
- Anterior-Posterior
- Transverse
- Checklist
- SP Junction
- Anterior-Posterior
- Transverse
- Checklist
- SSV Thigh Extension
- Anterior-Posterior
- Transverse
- Checklist
- AASV
- Anterior-Posterior
- Transverse
- Checklist
- PASV
- Anterior-Posterior
- Transverse
- Checklist
- Thigh Perf
- Anterior-Posterior
- Transverse
- Checklist
- Prox Calf Perf
- Anterior-Posterior
- Transverse
- Checklist
- Mid Calf Perf
- Anterior-Posterior
- Transverse
- Checklist
- Dist Calf Perf
- Anterior-Posterior
- Transverse
- Checklist
- Pseudoaneurysm
- Long
- Anterior-Posterior
- Transverse
- Neck
- D-Mode
- ACA
- A1 ACA
- MCA
- M1 MCA

- M2 MCA
- AComA
- Terminal ICA
- PComA
- PCA
- P1 PCA
- P2 PCA
- Ophthalmic A
- ICA Siphon
- Terminal Vert A
- BA
- Ba V
- CCA
- ICA
- ECA
- Bulb
- Carotid Bifurcation
- Vert A
- Subclav A
- Innom A
- Mammary A
- Subclav V
- CCA Aneurysm
- ICA Aneurysm
- ECA Aneurysm
- Bulb Aneurysm
- Carotid Bifurcation Aneurysm
- Vert A Aneurysm
- Subclav A Aneurysm
- Innom A Aneurysm
- Mammary A Aneurysm
- Carotid Graft 1 Native Inflow
- Carotid Graft 1 Anast Pre
- Carotid Graft 1 Anast Max
- Carotid Graft 1 Anast Post
- Carotid Graft 1 Graft
- Carotid Graft 1 Native Outflow
- Carotid Graft 2 Native Inflow
- Carotid Graft 2 Anast Pre
- Carotid Graft 2 Anast Max
- Carotid Graft 2 Anast Post
- Carotid Graft 2 Graft
- Carotid Graft 2 Native Outflow
- Carotid Graft 3 Native Inflow
- Carotid Graft 3 Anast Pre
- Carotid Graft 3 Anast Max
- Carotid Graft 3 Anast Post
- Carotid Graft 3 Graft
- Carotid Graft 3 Native Outflow
- Carotid Stent 1
- Carotid Stent 2
- Carotid Stent 3
- Axill A
- Brachial A
- Ulnar A
- Radial A
- UE A Graft 1 Native Inflow
- UE A Graft 1 Anast
- UE A Graft 1 Graft
- UE A Graft 1 Native Outflow
- UE A Graft 2 Native Inflow
- UE A Graft 2 Anast
- UE A Graft 2 Graft
- UE A Graft 2 Native Outflow
- UE A Graft 3 Native Inflow
- UE A Graft 3 Anast
- UE A Graft 3 Graft
- UE A Graft 3 Native Outflow
- UE A Stent 1
- UE A Stent 2
- UE A Stent 3
- C.Iliac A
- Ex.Iliac A
- IIA
- CFA
- DFA
- SFA
- Pop A
- TP Trunk A
- A.Tib A
- Peroneal A
- P.Tib A
- Dors.Ped. A
- LE A Graft 1 Native Inflow
- LE A Graft 1 Anast Pre
- LE A Graft 1 Anast Max
- LE A Graft 1 Anast Post
- LE A Graft 1 Graft
- LE A Graft 1 Native Outflow
- LE A Graft 2 Native Inflow
- LE A Graft 2 Anast Pre
- LE A Graft 2 Anast Max
- LE A Graft 2 Anast Post
- LE A Graft 2 Graft
- LE A Graft 2 Native Outflow
- LE A Graft 3 Native Inflow

- LE A Graft 3 Anast Pre
- LE A Graft 3 Anast Max
- LE A Graft 3 Anast Post
- LE A Graft 3 Graft
- LE A Graft 3 Native Outflow
- LE A Stent 1
- LE A Stent 2
- LE A Stent 3
- Axill V
- Brachial V
- Radial V
- Ulnar V
- Cephalic V
- Basilic V
- AVF-Inflow Artery
- AVF-Anast
- AVF-Outflow Vein Level 1
- AVF-Outflow Vein Level 2
- AVF-Outflow Vein Level 3
- AVF-Outflow Vein Level 4
- AVF-Outflow Vein Level 5
- AVF-Outflow Vein Level 6
- AVF-Stenosis 1
- AVF-Stenosis 2
- AVF-Stenosis 3
- AV Graft-Inflow Artery
- AV Graft-Arterial Anast
- AV Graft-Graft
- AV Graft-Venous Anast
- AV Graft-Outflow Vein Level 1
- AV Graft-Outflow Vein Level 2
- AV Graft-Outflow Vein Level 3
- AV Graft-Outflow Vein Level 4
- AV Graft-Outflow Vein Level 5
- AV Graft-Outflow Vein Level 6
- ASP
- BSP
- -----
- CCA(Sten)
- Pre Sten
- Sten
- Post Sten
- ICA(Sten)
- Pre Sten
- Sten
- Post Sten
- ECA(Sten)
- Pre Sten
- Sten
- Post Sten
- Carotid Bifurcation(Sten)
- Pre Sten
- Sten
- Post Sten
- Vert A(Sten)
- Pre Sten
- Sten
- Post Sten
- Subclav A(Sten)
- Pre Sten
- Sten
- Post Sten
- Innom A(Sten)
- Pre Sten
- Sten
- Post Sten
- Mammary A(Sten)
- Pre Sten
- Sten
- Post Sten
- Carotid Stenosis 1
- Pre Sten
- Sten
- Post Sten
- Carotid Stenosis 2
- Pre Sten
- Sten
- Post Sten
- Carotid Stenosis 3
- Pre Sten
- Sten
- Post Sten
- Carotid Stenosis 4
- Pre Sten
- Sten
- Post Sten
- Axill A(Sten)
- Pre Sten
- Sten
- Post Sten
- Brachial A(Sten)
- Pre Sten

- Sten
- Post Sten
- Ulnar A(Sten)
- Pre Sten
- Sten
- Post Sten
- Radial A(Sten)
- Pre Sten
- Sten
- Post Sten
- UE A Stenosis 1
- Pre Sten
- Sten
- Post Sten
- UE A Stenosis 2
- Pre Sten
- Sten
- Post Sten
- UE A Stenosis 3
- Pre Sten
- Sten
- Post Sten
- UE A Stenosis 4
- Pre Sten
- Sten
- Post Sten
- C.Iliac A(Sten)
- Pre Sten
- Sten
- Post Sten
- Ex.Iliac A(Sten)
- Pre Sten
- Sten
- Post Sten
- IIA(Sten)
- Pre Sten
- Sten
- Post Sten
- CFA(Sten)
- Pre Sten
- Sten
- Post Sten
- DFA(Sten)
- Pre Sten
- Sten
- Post Sten
- SFA(Sten)
- Pre Sten
- Sten
- Post Sten
- TP Trunk A(Sten)
- Pre Sten
- Sten
- Post Sten
- A.Tib A(Sten)
- Pre Sten
- Sten
- Post Sten
- Peroneal A(Sten)
- Pre Sten
- Sten
- Post Sten
- P.Tib A(Sten)
- Pre Sten
- Sten
- Post Sten
- Dors.Ped. A(Sten)
- Pre Sten
- Sten
- Post Sten
- LE A Stenosis 1
- Pre Sten
- Sten
- Post Sten
- LE A Stenosis 2
- Pre Sten
- Sten
- Post Sten
- LE A Stenosis 3
- Pre Sten
- Sten
- Post Sten
- LE A Stenosis 4
- Pre Sten
- Sten
- Post Sten
- C.Iliac V
- PV
- Reflux
- Checklist
- Ex.Iliac V
- PV

- Reflux
- Checklist
- IIV
- PV
- Reflux
- Checklist
- CFV
- PV
- Reflux
- Checklist
- Femoral V
- PV
- Reflux
- Checklist
- DFV
- PV
- Reflux
- Checklist
- Pop V
- PV
- Reflux
- Checklist
- P.Tib V
- PV
- Reflux
- Checklist
- Peroneal V
- PV
- Reflux
- Checklist
- Sural V
- PV
- Reflux
- Checklist
- Soleal V
- PV
- Reflux
- Checklist
- A.Tib V
- PV
- Reflux
- Checklist
- TP Trunk V
- PV
- Reflux
- Checklist
- Saph V
- PV
- Reflux
- Checklist
- SSV
- PV
- Reflux
- Checklist
- SF Junction
- PV
- Reflux
- Checklist
- GSV Thigh
- PV
- Reflux
- Checklist
- GSV Knee
- PV
- Reflux
- Checklist
- GSV Calf
- PV
- Reflux
- Checklist
- SP Junction
- PV
- Reflux
- Checklist
- SSV Thigh Extension
- PV
- Reflux
- Checklist
- AASV
- PV
- Reflux
- Checklist
- PASV
- PV
- Reflux
- Checklist
- Thigh Perf
- PV
- Reflux
- Checklist
- Prox Calf Perf
- PV
- Reflux
- Checklist
- Mid Calf Perf
- PV

- Reflux
- Checklist
- Dist Calf Perf
- PV
- Reflux
- Checklist
- ABI
- ASP
- BSP
- Small Parts
 - B-Mode
 - Thyroid L
 - Thyroid H
 - Thyroid W
 - Isthmus H
 - THY Nodule1 Strain
 - THY Nodule2 Strain
 - THY Nodule3 Strain
 - THY Nodule1 Elas.
 - THY Nodule2 Elas.
 - THY Nodule3 Elas.
 - Breast Mass1 Strain
 - Breast Mass1 Elas.
 - Breast Mass2 Strain
 - Breast Mass2 Elas.
 - Breast Mass3 Strain
 - Breast Mass3 Elas.
 - Breast Mass4 Strain
 - Breast Mass4 Elas.
 - Breast Mass5 Strain
 - Breast Mass5 Elas.
 - Breast Mass6 Strain
 - Breast Mass6 Elas.
 - Breast Mass7 Strain
 - Breast Mass7 Elas.
 - Breast Mass8 Strain
 - Breast Mass8 Elas.
 - Breast Mass9 Strain
 - Breast Mass9 Elas.
 - Breast Mass10 Strain
 - Breast Mass10 Elas.
 - Testicular L
 - Testicular H
 - Testicular W
 - Epididymis L
 - Epididymis H
 - Epididymis W
 - Scrotal Wall
 - Testis V(2D)
 - Testis V(Valsalva 2D)
 - -----
 - Thyroid Vol
 - Testicular Vol
 - -----
 - Thyroid(Superior)
 - Anterior-Posterior
 - Transverse
 - Thyroid(Mid)
 - Anterior-Posterior
 - Transverse
 - Thyroid(Inferior)
 - Anterior-Posterior
 - Transverse
 - Parathyroid 1
 - Long
 - Anterior-Posterior
 - Transverse
 - Parathyroid 2
 - Long
 - Anterior-Posterior
 - Transverse
 - Parotid
 - Long
 - Anterior-Posterior
 - Transverse
 - Lymph Node 1
 - Long
 - Anterior-Posterior
 - Transverse
 - Lymph Node 2
 - Long
 - Anterior-Posterior
 - Transverse
 - Lymph Node 3
 - Long
 - Anterior-Posterior
 - Transverse
 - Lymph Node 4
 - Long
 - Anterior-Posterior
 - Transverse
 - Lymph Node 5
 - Long
 - Anterior-Posterior
 - Transverse
 - Lymph Node 6

-	Long	-	A
-	Anterior-Posterior	-	B
-	Transverse	-	THY Nodule2 Strain Ratio
-	Thyroid	-	A
-	Thyroid L	-	B
-	Thyroid H	-	THY Nodule3 Strain Ratio
-	Thyroid W	-	A
-	Thyroid Nodule 1	-	B
-	Length	-	THY Nodule1 Elas. Ratio
-	Height	-	A
-	Width	-	B
-	Thyroid Nodule 2	-	THY Nodule2 Elas. Ratio
-	Length	-	A
-	Height	-	B
-	Width	-	THY Nodule3 Elas. Ratio
-	Thyroid Nodule 3	-	A
-	Length	-	B
-	Height	-	Breast Mass 1
-	Width	-	L
-	Thyroid Nodule 4	-	H
-	Length	-	W
-	Height	-	Nip. Dist.
-	Width	-	Skin Dist.
-	Thyroid Nodule 5	-	Breast Mass 2
-	Length	-	L
-	Height	-	H
-	Width	-	W
-	Thyroid Nodule 6	-	Nip. Dist.
-	Length	-	Skin Dist.
-	Height	-	Breast Mass 3
-	Width	-	L
-	Thyroid Nodule 7	-	H
-	Length	-	W
-	Height	-	Nip. Dist.
-	Width	-	Skin Dist.
-	Thyroid Nodule 8	-	Breast Mass 4
-	Length	-	L
-	Height	-	H
-	Width	-	W
-	Thyroid Nodule 9	-	Nip. Dist.
-	Length	-	Skin Dist.
-	Height	-	Breast Mass 5
-	Width	-	L
-	Thyroid Nodule 10	-	H
-	Length	-	W
-	Height	-	Nip. Dist.
-	Width	-	Skin Dist.
-	THY Nodule1 Strain Ratio	-	Breast Mass 6

-	L	-	A
-	H	-	B
-	W	-	Breast Mass4 Elas. Ratio
-	Nip. Dist.	-	A
-	Skin Dist.	-	B
-	Breast Mass 7	-	Breast Mass5 Strain Ratio
-	L	-	A
-	H	-	B
-	W	-	Breast Mass5 Elas. Ratio
-	Nip. Dist.	-	A
-	Skin Dist.	-	B
-	Breast Mass 8	-	Breast Mass6 Strain Ratio
-	L	-	A
-	H	-	B
-	W	-	Breast Mass6 Elas. Ratio
-	Nip. Dist.	-	A
-	Skin Dist.	-	B
-	Breast Mass 9	-	Breast Mass7 Strain Ratio
-	L	-	A
-	H	-	B
-	W	-	Breast Mass7 Elas. Ratio
-	Nip. Dist.	-	A
-	Skin Dist.	-	B
-	Breast Mass 10	-	Breast Mass8 Strain Ratio
-	L	-	A
-	H	-	B
-	W	-	Breast Mass8 Elas. Ratio
-	Nip. Dist.	-	A
-	Skin Dist.	-	B
-	Breast Mass1 Strain Ratio	-	Breast Mass9 Strain Ratio
-	A	-	A
-	B	-	B
-	Breast Mass1 Elas. Ratio	-	Breast Mass9 Elas. Ratio
-	A	-	A
-	B	-	B
-	Breast Mass2 Strain Ratio	-	Breast Mass10 Strain Ratio
-	A	-	A
-	B	-	B
-	Breast Mass2 Elas. Ratio	-	Breast Mass10 Elas. Ratio
-	A	-	A
-	B	-	B
-	Breast Mass3 Strain Ratio	-	Testis
-	A	-	Testicular L
-	B	-	Testicular H
-	Breast Mass3 Elas. Ratio	-	Testicular W
-	A	-	Testis Mass 1
-	B	-	d1
-	Breast Mass4 Strain Ratio	-	d2

- d3
- Testis Mass 2
- d1
- d2
- d3
- Testis Mass 3
- d1
- d2
- d3
- Epididymis
- Epididymis L
- Epididymis H
- Epididymis W
- Testicle(Superior)
- H
- W
- Testicle(Mid)
- H
- W
- Testicle(Inferior)
- H
- W
- Epididymal Head
- L
- H
- W
- Epididymal Body
- L
- H
- W
- Epididymal Tail
- L
- H
- W
- D-Mode
- STA
- ITA
- Isthmus
- Parathyroid 1
- Parathyroid 2
- Testis A
- Testis V
- Testis V(Valsalva)
- Epididymis A
- Epididymis V
- Emergency
- B-Mode
- Renal L
- Renal H
- Renal W
- CBD
- Portal V Diam
- CHD
- GB wall th
- Aorta Bif
- Ureter
- Pre-BL L
- Pre-BL H
- Pre-BL W
- Post-BL L
- Post-BL H
- Post-BL W
- GS
- YS L
- CRL
- BPD
- UT L
- UT H
- UT W
- Endo
- Ovary L
- Ovary H
- Ovary W
- -----
- Renal Vol
- Pre-BL Vol
- Post-BL Vol
- Mictur.Vol
- Ovary Vol
- UT Vol
- UT SUM
- -----
- Uterus
- UTL
- UTH
- UTW
- Endo
- Ovary
- Ovary L
- Ovary H
- Ovary W
- Kidney
- Renal L
- Renal H
- Renal W
- Cortex

- Bladder
 - Pre-BL L
 - Pre-BL H
 - Pre-BL W
 - Post-BL L
 - Post-BL H
 - Post-BL W
- M-Mode**
- FHR (M)
- D-Mode**
- FHR (Doppler)
- Orthopedic
- B-Mode**
- HIP
 - HIP(α)
 - HIP(β)
 - d/D